

Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq

Date of Issue: January 8, 2025

Page 1 of 5

REQUEST NUMBER	: By Mail	APPROVED BY LAB INCHARGE QC
JOB NUMBER	: QC-B-01-25-0014	
CERTIFICATE NUMBER	: QC-B-01-25-0014-01	



CUSTOMER DETAILS

Name : Egyptian Maintenance Company(EMC) IRAQ
Address : Iraq, Basra

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : True RMS Multimeter
Type of Indication : Digital
Manufacturer : FLUKE
Model : 115
Serial Number : 24691655

Calibrated Range :

Voltage (DC)	0.01 mV	to	600 V
Voltage (AC) @ 50 Hz	0.1 mV	to	600 V
Voltage (AC) @ 1 kHz	0.1 mV	to	600 V
Current (DC)	0.01 μA	to	10 A
Current (AC) @ 50 Hz	0.1 μA	to	10 A
Current (AC) @ 1 kHz	0.1 μA	to	10 A
Resistance	1 Ω	to	6 M Ω

Resolution :

Voltage (DC)	0.01 mV	0.001 mV	0.001 V
Voltage (AC) @ 50 Hz	0.1 mV	0.001 mV	0.001 V
Voltage (AC) @ 1 kHz	0.1 mV	0.001 mV	0.001 V
Current (DC)	0.01 μA	0.001 mA	0.001 A
Current (AC) @ 50 Hz	0.1 μA	0.001 mA	0.001 A
Current (AC) @ 1 kHz	0.1 μA	0.001 mA	0.001 A
Resistance	0.1 Ω	0.001 K Ω	0.001 M Ω

Calibrated By : Hussein Alaa
Calibration Date : January 8, 2025
Calibration Due : January 7, 2026

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C ± 2 °C
Relative Humidity : 45 %RH ± 5 %RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with QC Calibration Procedure # QC/CP/E/01

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Multifunction Calibrator	Fluke, USA	5522A	2806902	November 5, 2024	November 4, 2025
Ref Multimeter	Fluke, USA	8508A	276568089	November 5, 2024	November 4, 2025
Decade Resistance Box	Corpico	RBB6-B	18F-1093	November 5, 2024	November 4, 2025
Programmable Inductance substitutor	IET Labs	PLS-1492	J1-1419517	November 5, 2024	November 4, 2025

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-B-01-25-0014-01

Page 2 of 5

REQUEST NUMBER: By Email
JOB NUMBER: QC-B-01-25-0014

CALIBRATION TEST RESULTS

Measurement Data for DC Voltage Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μV	μV
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
μV	μV	μV	$\pm (\mu\text{V})$
99.9	100.00	-0.10	0.1
mV	mV	mV	$\pm (\text{mV})$
0.9	0.98	-0.08	0.1
9.8	9.99	-0.19	0.1
100	100.03	-0.03	0.1
500	500.07	-0.07	0.1
600	601.79	-1.79	0.1
V	V	V	$\pm (\text{V})$
1.002	1.00	0.00	0.1
-1.002	-1.00	0.00	0.1
10.01	10.00	0.01	0.1
100	99.98	0.02	0.1
300	301.85	-1.85	0.1
500	501.06	-1.06	1
600	602.00	-2.00	1

Measurement Data for AC Voltage @ 50 Hz Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μV	μV
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
μV	μV	μV	$\pm (\mu\text{V})$
99.9	100.000	-0.100	0.1
mV	mV	mV	$\pm (\text{mV})$
0.9	1.000	-0.100	0.1
9.8	10.000	-0.200	0.1
100	100.000	0.000	0.1
500	499.998	0.002	0.1
600	598.650	1.350	0.1
V	V	V	$\pm (\text{V})$
1.002	0.998400	0.003600	0.1
10.01	9.99632	0.01368	0.1
100.0	100.0147	-0.0147	0.1
300.0	300.013	-0.013	0.1
500	500.015	-0.015	1
600	600.942	-0.942	1

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-B-01-25-0014-01

Page 3 of 5

REQUEST NUMBER: By Email
JOB NUMBER: QC-B-01-25-0014

Measurement Data for AC Voltage @ 1 kHz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μV	μV
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μV	μV	μV	$\pm (\mu V)$
99.9	100.00	-0.10	0.1
mV	mV	mV	$\pm (mV)$
0.9	1.001	-0.101	0.1
9.8	10.001	-0.201	0.1
100	100.001	-0.001	0.1
500	500.000	0.000	0.1
600	599.950	0.050	0.1
V	V	V	$\pm (V)$
1.002	1.000010	0.001990	0.1
10.01	10.00001	0.00999	0.1
100.0	99.9870	0.0130	0.1
300.0	299.932	0.068	0.1
500	499.974	0.026	1
600	600.989	-0.989	1

Measurement Data for DC Current

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μA	μA	μA	$\pm (\mu A)$
99.9	99.999	-0.099	0.1
mA	mA	mA	$\pm (mA)$
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	99.985	0.015	0.1
500	499.994	0.006	0.1
600	600.125	-0.125	0.1
A	A	A	$\pm (A)$
1.002	1.000010	0.001990	0.01
-1.002	-1.000030	-0.001970	0.01
3.002	2.994550	0.007450	0.01
5.003	4.999950	0.003050	0.06
10.003	9.999900	0.003100	0.06

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-B-01-25-0014-01

Page 4 of 5

REQUEST NUMBER:

By Email

JOB NUMBER:

QC-B-01-25-0014-01

Measurement Data for AC Current @ 50 Hz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
μA	μA	μA	$\pm (\mu\text{A})$
99.9	99.999	-0.099	0.1
mA	mA	mA	$\pm (\text{mA})$
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	100.000	0.000	0.1
500	500.030	-0.030	0.1
900	899.900	0.100	0.2
A	A	A	$\pm (\text{A})$
1.002	1.000100	0.001900	0.01
3.002	3.000040	0.001960	0.01
5.001	5.008040	-0.007040	0.06
9.9	10.235000	-0.335000	0.06

Measurement Data for AC Current @ 1 kHz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Measurement Data for AC Current @ 1 kHz

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
μA	μA	μA	$\pm (\mu\text{A})$
99.9	99.999	-0.099	0.1
mA	mA	mA	$\pm (\text{mA})$
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	100.020	-0.020	0.1
500	500.040	-0.040	0.1
600	601.230	-1.230	0.2
A	A	A	$\pm (\text{A})$
1.002	0.999950	0.002050	0.01
3.002	2.999780	0.002220	0.01
5.002	4.999780	0.002220	0.06
10.001	10.002568	-0.001568	0.06

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-B-01-25-0014-01

Page 5 of 5

REQUEST NUMBER:

By Email

JOB NUMBER:

QC-B-01-25-0014

Measurement Data for Resistance

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
Ω	Ω
0	0

Before Adjustment	After Adjustment
$k\Omega$	$k\Omega$
0	0

Before Adjustment	After Adjustment
$M\Omega$	$M\Omega$
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
Ω	Ω	Ω	$\pm (\Omega)$
1.1	1.00010	0.09990	0.08
10.1	9.99990	0.10010	0.10
100.1	99.9918	0.1082	0.08
900.2	900.0000	0.2000	0.08
$k\Omega$	$k\Omega$	$k\Omega$	$\pm (k\Omega)$
1.002	1.000000	0.002000	0.01
10.002	9.99876	0.00324	0.01
100.1	100.0006	0.0994	0.06
500.1	499.9993	0.1007	0.05
900.1	899.999	0.1010	0.08
$M\Omega$	$M\Omega$	$M\Omega$	$\pm (M\Omega)$
1.001	1.000001	0.000999	0.12
2.000	2.00420	-0.00420	0.01
4.001	3.9970	0.0040	0.02
6.003	5.9450	0.0580	0.02

Results :

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- The results are as found (no adjustment done).
- The results are post adjustment.

This certificate is issued in accordance with the laboratory accreditation requirements as per ISO/IEC17025:2019.

This certificate may not be reproduced other than in full, unless permission for the publication of an approved extract has been obtained in writing from the director of QC. It does not of itself impute to the subject of the calibration any attributes beyond those shown by the data contained herein.