

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: March 13, 2025

Page 1 of 1

REQUEST NUMBER : By Contract JOB NUMBER : QC-CAL-25094 CERTIFICATE NUMBER : QC-CAL-25094-21	APPROVED BY LAB INCHARGE QC  Asjad Rana
--	--

CUSTOMER DETAILS

Name : Honghua Oil & Gas Engineering Services Limited
Address : Basra, South Iraq
Department : HH008



EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Gauge
Manufacturer : WIKA
Model : EN837-1
Serial Number : HH14042502
Location : Cascade System
Type : Analog
Range : 0 to 300 psi
Calibrated Range : 0 to 300 psi
Resolution : 10 psi
Accuracy Class : 1.6
Tolerance : 1.6 % of Full Scale
As found : In Tolerance
Calibrated By : Abdulrahman Loay
Calibration Date : March 13, 2025
Calibration Due : **March 12, 2026**



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C **Humidity** : 44.0 ± 5 % RH **Atm. Pressure** : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±1
0	0	0.00	0.0
50.0	50.05	-0.05	-0.02
150.0	150.07	-0.07	-0.02
225.0	225.10	-0.10	-0.03
300.0	300.50	-0.50	-0.17

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.