

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** January 17, 2025

**Page 1 of 1**

REQUEST NUMBER	: By Contract	CUSTOMER DETAILS
JOB NUMBER	: QC-CAL-25021	Name : <b>Halliburton Worldwide IRAQ</b> (CMT)
CERTIFICATE NUMBER	: <b>QC-CAL-25021-15</b>	Address : Western Burjesia, Oil Street, District Zubair-South Iraq

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Relief Valve**  
 Manufacturer : Kunkle Valve  
 Model : 6283FEM01AKM  
 Serial Number : 634810.10.10  
 Set/Pressure : 36 psi  
 Tolerance : ± 2 psi  
 As Found : In Tolerance  
 Calibration Date : January 17, 2025  
**Calibration Due : 16 January 2027**  
 Last Calibration : NA



### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C      Humidity : 30.0 ± 5 % RH      Atm. Pressure : 1011.0 ± 10 hPa

### CALIBRATION METHOD

Standard Method API 527 was used to test the unit under test. Test pressure was applied on inlet of unit under test and test was repeated 3 times. Identical repetitions were carried out as recommended in used standard.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	1505119397	Sino Instruments / HS108	05 Nov, 2024	04 Nov, 2025

### CALIBRATION TEST RESULTS

Set Pressure on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
psi	psi	psi	± psi
36	35.8	0.2	2.0

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

**DEVIATION FROM STANDARD METHOD :** None

**REMARK (S) :**

- The results are as found (no adjustment done).  
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY / LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE      Asjad Rafiq	