

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** January 29, 2025

**Page 1 of 1**

<b>REQUEST NUMBER</b> : By Mail <b>JOB NUMBER</b> : QC-CAL-25038 <b>CERTIFICATE NUMBER</b> : <b>QC-CAL-25038-11</b>	<b>CUSTOMER DETAILS</b> <b>Name</b> : <b>Halliburton Worldwide IRAQ</b> (Sperry) <b>Address</b> : Western Burjesia, Oil Street, District Zubair-South Iraq
---	--

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

<b>Description</b> : <b>Pressure Transducer</b> <b>Manufacturer</b> : Hohner <b>Model</b> : IP66 <b>Serial Number</b> : 101121-39 <b>SAP</b> : 30039995 <b>Range</b> : 0 to 10000 PSI <b>Output</b> : 4 to 20 mA <b>As Found</b> : Good	<b>Calibration Date</b> : January 29, 2025 <b>Calibration Due</b> : <b>January 28, 2026</b>
--	--



### ENVIRONMENTAL CONDITIONS DURING TEST

**Ambient Temperature** : 22.5 ± 1°C    **Humidity** : 30.0 ± 5 % RH    **Atm. Pressure** : 1013.6 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard and measuring mA at the output of Pressure Transducer by using Process Calibrator having traceability to Internationally recognised standards.

### TRACEABILITY

The measurements made by Quality Control Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

### REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	Make / Model	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5-Nov-24	4-Nov-25
Process Calibrator	22564482	FLUKE / 725	6-Nov-24	5-Nov-25

### CALIBRATION TEST RESULTS

Ref. Standard Values (A)-INPUT		Measured Values on Unit Under Calibration (M)-OUTPUT		Deviation (D=M-A)		Expanded Uncertainty (U)
PSI	% of Range	PSI	mA	PSI	mA	±mA
0	0%	0.00	3.999	0.00	-0.001	0.10
2500	25%	2502.00	7.998	2.00	-0.002	0.10
5000	50%	5003.00	11.998	3.00	-0.002	0.10
7500	75%	7503.00	15.997	3.00	-0.003	0.10
10000	100%	10002.00	19.998	2.00	-0.002	0.10

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

CALIBRATED BY	REVIEWED & APPROVED BY / LAB INCHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE      Asjad Rafiq	