

## Certificate of Calibration

AlTakamul Yard, North Rumaila-South Iraq

● Phone : +964 7810009138 ● www.qualitycontrol-iraq.com ● E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** April 15, 2025

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY LAB INCHARGE QC
JOB NUMBER	: QC-CAL-25132	
CERTIFICATE NUMBER	: QC-CAL-25132-01	



### CUSTOMER DETAILS

Name : **Honghua Oil & Gas Engineering Services Limited**  
 Address : Basra, South Iraq  
 Rig : HH008

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Multigas Detector**  
 Manufacturer : Honeywell  
 Model : BW Max XT II  
 Serial Number : MA220-007356  
 Type : Digital  
 Calibrated Range : 25 PPM H2S      100 PPM CO  
                                  18% O2                      50% LEL

Alarm Details	Low	High
Oxygen (O2) %	19.5	23.5
H2S-ppm	10	20
CH4-%LEL	10	20
CO-ppm	35	200

Calibrated By : Hussein Alaa  
 Calibration Date : April 15, 2025  
**Calibration Due : October 14, 2025**



### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25 °C ± 2 °C  
 Relative Humidity : 35 %RH ± 5 %RH

### CALIBRATION METHOD

A reference cylinder consisting of composition gases (O2, LEL, H2S, CO) is used, which are detected by gas detector by diffusion.

### TRACEABILITY

The measurements made by Aberdeen Calibration & Testing Labs are traceable to NIST, vide CALGAZ Certificate of Analysis  
 The measurements made by Aberdeen Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by Aberdeen.

### REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE	GAS EXPIRY DATE
Mixture of Gases.	950-092876	CALGAZ	July 31, 2026

### CALIBRATION TEST RESULTS

Gas Name	Values of UUC	Values on Reference Standard	Error in Reading of UUC	Uncertainty of Measurement
Oxygen(O2)%	18	18	0	± 1
Methane(CH4) LEL	50	50	0	± 1
Carbon Monoxide(CO) ppm	100	100	0	± 1
Hydrogen Sulphide(H2S) ppm	25	25.0	0	± 1

### Results :

The above Value Indicates That the Instrument complies with the Specified Maximum Permissible error tolerance.  
 At The Time of Calibration. and ready for use.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

**DEVIATION FROM STANDARD METHOD :** None

### REMARK (S) :

- The results are as found (no adjustment done).
- The results are post adjustment.