

## Certificate of Calibration

AlTakamul Yard, North Rumaila-South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com


Page 1 of 1

Date of Issue:

May 3, 2025

REQUEST NUMBER : By Contract  
JOB NUMBER : QC-CAL-25155  
CERTIFICATE NUMBER : QC-CAL-25155-02

APPROVED BY LAB INCHARGE-QC



**CUSTOMER DETAILS**

Name : **Daqing Drilling Company Iraq**  
Address : South Rumaila South Iraq  
Rig : DQ1202

**EQUIPMENT IDENTIFICATION AND SPECIFICATIONS**

Description : **H2S Portable Sensor**  
Manufacturer : HONEYWELL  
Model : BW CLIP H2S  
Serial Number : 5220BWC01241402247  
Type : Digital  
Calibrated Range : 25 PPM H2S  
As found : Good  
Calibrated By : Hussein Alaa  
Calibration Date : May 3, 2025  
**Calibration Due : August 2, 2025**



Alarm Details	Low	High
H2S PPM	10	15

**ENVIRONMENTAL CONDITIONS DURING TEST**

Ambient Temperature : 22 °C ± 2 °C  
Relative Humidity : 30 %RH ± 5 %RH

**CALIBRATION METHOD**

A reference cylinder consisting of Sulpher Dioxide (H2S) Gas is used , which are detected by gas detector by diffusion.

**TRACEABILITY**

The measurements made by QC Calibration & Testing Labs are traceable to NIST, vide Calgaz Certificate of Analysis  
The measurements made by QC Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

**REFERENCE EQUIPMENT USED :**

DESCRIPTION	SERIAL NO.	MAKE	GAS EXPIRY DATE
Multigas Calibration cylinder (25 PPM H2S)	WO463909-3	CALGAZ	July 31, 2026

**CALIBRATION TEST RESULTS**

Gas Name	Values of UUC	Values on Reference Standard	Error in Reading of UUC	Uncertainty of Measurement
H2S Gas (25 ppm)	25.0	25	0	± 0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.