

## Certificate of Calibration

AlTakamul Yard, North Rumaila-South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:**

April 4, 2025

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY LAB INCHARGE-QC
JOB NUMBER : QC-CAL-25157	
CERTIFICATE NUMBER : QC-CAL-25157-01	

**CUSTOMER DETAILS**

Name : **Daqing Drilling Company Iraq**  
 Address : South Rumaila South Iraq  
 Rig : DQ1202

**EQUIPMENT IDENTIFICATION AND SPECIFICATIONS**

Description : **H2S Portable Sensor**  
 Manufacturer : HONEYWELL  
 Model : BW CLIP H2S  
 Serial Number : 5220BWC01224004907  
 Type : Digital  
 Calibrated Range : 25 PPM H2S  
 As found : Good  
 Calibrated By : Hussein Alaa  
 Calibration Date : April 4, 2025  
**Calibration Due : July 3, 2025**



Alarm Details	Low	High
H2S PPM	10	15

**ENVIRONMENTAL CONDITIONS DURING TEST**

Ambient Temperature : 22 °C ± 2 °C  
 Relative Humidity : 30 %RH ± 5 %RH

**CALIBRATION METHOD**

A reference cylinder consisting of Sulphur Dioxide (H2S) Gas is used , which are detected by gas detector by diffusion.

**TRACEABILITY**

The measurements made by QC Calibration & Testing Labs are traceable to NIST, vide Calgaz Certificate of Analysis  
 The measurements made by QC Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

**REFERENCE EQUIPMENT USED :**

DESCRIPTION	SERIAL NO.	MAKE	GAS EXPIRY DATE
Multigas Calibration cylinder (25 PPM H2S)	WO463909-3	CALGAZ	July 31, 2026

**CALIBRATION TEST RESULTS**

Gas Name	Values of UUC	Values on Reference Standard	Error in Reading of UUC	Uncertainty of Measurement
H2S Gas (25 ppm)	25.0	25	0	± 0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.