

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 19, 2025

Page 1 of 1

REQUEST NUMBER : By Mail JOB NUMBER : QC-CAL-25276 CERTIFICATE NUMBER : QC-CAL-25276-01	CUSTOMER DETAILS Name : Halliburton Worldwide IRAQ (Sperry) Address : Western Burjesia, Oil Street, District Zubair-South Iraq
---	---

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Dial Indicator Gauge
Manufacturer : Starrett
Model : 25-441
Serial Number : DIL-12
SAP No : 300098405
Type : Analog
Range : 0 to 1 inch
Calibration Range : 0 to 1 inch
Readability : 0.001"
As Found : In Tolerance
Calibration Date : July 19, 2025
Calibration Due : **18 July 2026**
Last Calibration : April 22, 2024



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 30.0 ± 5 % RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.10M-2001

The reported uncertainty is the expanded uncertainty with k=2 according to the

ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

REFERENCE EQUIPMENT USED :


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 5, 2024	November 4, 2025

CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
0.000	0.000	0.000	0.000
0.200	0.200	0.000	0.001
0.500	0.500	0.000	0.001
0.800	0.800	0.000	0.001
1.000	1.000	0.000	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.

CALIBRATED BY	REVISED & APPROVED BY IN CHARGE	CLIENT
 Abdulrahman Loay	 LAB INCHARGE Asjad Rafiq	

