

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 19, 2025

Page 1 of 1

<b>REQUEST NUMBER</b> : By Mail <b>JOB NUMBER</b> : QC-CAL-25276 <b>CERTIFICATE NUMBER</b> : <b>QC-CAL-25276-02</b>	<b>CUSTOMER DETAILS</b> <b>Name</b> : Halliburton Worldwide IRAQ (Sperry) <b>Address</b> : Western Burjesia, Oil Street, District Zubair-South Iraq
---	---

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

**Description** : Dial Indicator Gauge  
**Manufacturer** : Starrett  
**Model** : 25-441  
**Serial Number** : DIL-26  
**SAP No** : 300179172  
**Type** : Analog  
**Range** : 0 to 1 inch  
**Calibration Range** : 0 to 1 inch  
**Readability** : 0.001"  
**As Found** : In Tolerance  
**Calibration Date** : July 19, 2025  
**Calibration Due** : **18 July 2026**  
**Last Calibration** : July 8, 2024



### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 30.0 ± 5 % RH

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.10M-2001

The reported uncertainty is the expanded uncertainty with k=2 according to the

ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

### REFERENCE EQUIPMENT USED :


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 5, 2024	November 4, 2025

### CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
0.000	0.000	0.000	0.000
0.200	0.200	0.000	0.001
0.500	0.500	0.000	0.001
0.800	0.800	0.000	0.001
1.000	0.999	-0.001	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.

<b>CALIBRATED BY</b>  Abdulrahman Loay	<b>REVIEWED &amp; APPROVED BY</b>  ASJAD RAFIQ <b>LAB INCHARGE</b>	<b>CLIENT</b>
---	--	---------------

