

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 19, 2025

Page 1 of 1

<b>REQUEST NUMBER</b> : By Mail <b>JOB NUMBER</b> : QC-CAL-25276 <b>CERTIFICATE NUMBER</b> : <b>QC-CAL-25276-03</b>	<b>CUSTOMER DETAILS</b> Name : <b>Halliburton Worldwide IRAQ</b> (Sperry) Address : Western Burjesia, Oil Street, District Zubair-South Iraq
---	--

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Dial Indicator Gauge**  
 Manufacturer : Starrett  
 Model : 25-441  
 Serial Number : DIL-30  
 SAP No : 300254146  
 Type : Analog  
 Range : 0 to 1 inch  
 Calibration Range : 0 to 1 inch  
 Readability : 0.001"  
 As Found : In Tolerance  
 Calibration Date : July 19, 2025  
**Calibration Due : 18 July 2026**  
 Last Calibration : July 16, 2024



### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 30.0 ± 5 % RH

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.10M-2001  
 The reported uncertainty is the expanded uncertainty with k=2 according to the  
 ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

### REFERENCE EQUIPMENT USED :


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 5, 2024	November 4, 2025

### CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
0.000	0.000	0.000	0.000
0.200	0.200	0.000	0.001
0.500	0.500	0.000	0.001
0.800	0.799	-0.001	0.001
1.000	0.999	-0.001	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.

CALIBRATED BY   Abdulrahman Loay	REVISED & APPROVED BY INCHARGE   LAB INCHARGE Asjad Rafiq	CLIENT
--	---	--------