

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 14, 2025

Page 1 of 1

REQUEST NUMBER	: By Mail	CUSTOMER DETAILS	
JOB NUMBER	: QC-CAL-25266	Name	Halliburton Worldwide IRAQ (HCT)
CERTIFICATE NUMBER	: QC-CAL-25266-15	Address	Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : Resato
 Model : EN837-1
 Serial Number : **1108359794**
 Type : Analog
 Location : Pressure Pump SN. 37042
 Range : 0 to 23000 psi
 Calibrated Range : 0 to 23000 psi
 Resolution : 500 psi
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : July 14, 2025
Calibration Due : July 13, 2026



DEVIATION FROM STANDARD METHOD : None
REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 29.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
0	0	0	%
0	0.00	0.00	0.00
6000.0	6014.0	-14.00	-0.06
12000.0	12028.0	-28.00	-0.12
18000.0	18042.0	-42.00	-0.18
23000.0	23056.0	-56.00	-0.24

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIEWED & LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	