

# Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq

Date of Issue: July 14, 2025

Page 1 of 5

REQUEST NUMBER : By Mail	APPROVED BY: [Signature]
JOB NUMBER : QC-CAL-25266	APPROVED FOR: [Signature]
CERTIFICATE NUMBER : QC-CAL-25266-20	

### CUSTOMER DETAILS

Name : Halliburton Worldwide IRAQ  
Department : HCT  
Address : Western Burjesia, Oil Street, Zubair, South Iraq

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Multimeter**  
Type of Indication : Digital  
Manufacturer : Fluke, USA  
Model : 179  
Serial Number : 20700064



#### Calibrated Range:

Voltage (DC)	0.1 mV	to	1000 V
Voltage (AC) @ 50 Hz	1 mV	to	1000 V
Voltage (AC) @ 1 kHz	1 mV	to	1000 V
Current (DC)	0.01 mA	to	10 A
Current (AC) @ 50 Hz	0.01 mA	to	10 A
Current (AC) @ 1 kHz	0.01 mA	to	10 A
Resistance	0.1 ohm	to	60 M-ohm

#### Resolution:

Voltage (DC)	0.1 mV	0.001 V	0.01V	0.1V	1V
Voltage (AC) @ 50 Hz	0.1 mV	0.001 V	0.01V	0.1V	1V
Voltage (AC) @ 1 kHz	0.1 mV	0.001 V	0.01V	0.1V	1V
Current (DC)	0.01 mA	0.1mA	0.001A	0.01A	
Current (AC) @ 50 Hz	0.01 mA	0.1mA	0.001A	0.01A	
Current (AC) @ 1 kHz	0.01 mA	0.1mA	0.001A	0.01A	
Resistance	0.1 ohm	0.001 k-ohm	0.01k-ohm	0.1k-ohm	0.001M-ohm 0.01M-ohm

As Found : In Tolerance  
Calibration Date : July 14, 2025  
Calibration Due : July 13, 2026  
Last Calibration : NA

Ambient Temperature : 22 °C ± 2 °C  
Relative Humidity : 40 %RH ± 5 %RH

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with QC Calibration Procedure # QC/CP/E/01

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

### TRACEABILITY

The measurements made by Quality Control Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

### REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Multifunction Calibrator	Fluke, USA	5522A	2806902	8/20/2024	8/19/2025
Ref Multimeter	Fluke, USA	8508A	276568089	8/20/2024	8/19/2025
Decade Resistance Box	Corpico	RBB6-B	18F-1093	8/20/2024	8/19/2025
Programmable Inductance substitutor	IET Labs	PLS-1492	J1-1419517	8/20/2024	8/19/2025

## CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC-CAL-25266-20

Page 2 of 5

REQUEST NUMBER: By Mail  
JOB NUMBER: QC-CAL-25266

### CALIBRATION TEST RESULTS

#### Measurement Data for DC Voltage

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu\text{V}$	$\mu\text{V}$
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
mV	mV	mV	$\pm$ (mV)
1.0	1.00	0.00	0.1
9.9	10.00	-0.10	0.1
101.2	100.00	1.20	0.1
502.4	500.00	2.40	0.1
599.8	600.00	-0.20	0.1
V	V	V	$\pm$ (V)
1.003	1.00	0.00	0.1
5.002	5.00	0.00	0.1
10.01	10.00	0.01	0.1
100	100.00	0.00	0.1
301	300.00	1.00	0.1
501	500.00	1.00	0.2
898	900.00	-2.00	0.2
998	1000.00	-2.00	0.2

#### Measurement Data for AC Voltage @ 50 Hz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu\text{V}$	$\mu\text{V}$
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
mV	mV	mV	$\pm$ (mV)
0.9	1.000	-0.100	0.1
9.9	10.000	-0.100	0.1
100.0	100.000	0.000	0.1
500.8	499.998	0.802	0.1
600.2	600.000	0.200	0.1
V	V	V	$\pm$ (V)
1.002	0.999980	0.002020	0.1
10.01	9.99998	0.01002	0.1
100.0	100.0130	-0.0130	0.1
300.0	300.013	-0.013	0.1
500	500.013	-0.013	0.1
901	900.080	0.920	0.2
1000	1000.056	-0.056	0.2

**CERTIFICATE OF CALIBRATION**

CERTIFICATE NUMBER

QC-CAL-25266-20

REQUEST NUMBER: By Mail  
JOB NUMBER: QC-CAL-25266

Page 3 of 5

**Measurement Data for AC Voltage @ 1 kHz**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu V$	$\mu V$
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
mV	mV	mV	$\pm$ (mV)
0.9	1.001	-0.101	0.1
9.8	10.001	-0.201	0.1
100.0	100.003	-0.003	0.1
500.0	500.000	0.000	0.1
600.0	600.000	0.000	0.1
V	V	V	$\pm$ (V)
1.002	1.000010	0.001990	0.1
10.01	10.00001	0.00999	0.1
100.0	99.9870	0.01300	0.1
300.0	299.987	0.0130	0.1
500	499.987	0.0130	0.2
901	899.930	1.0700	0.2
1000	1000.001	-0.0010	0.3

**Measurement Data for DC Current**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu A$	$\mu A$
0	0

Before Adjustment	After Adjustment
$\mu A$	$\mu A$
0	0

Before Adjustment	After Adjustment
$\mu A$	$\mu A$
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
mA	mA	mA	$\pm$ (mA)
0.90	1.000	-0.100	0
9.80	10.000	-0.200	0.1
100.0	99.999	0.001	0.1
500.0	499.994	0.006	0.1
599.5	600.000	-0.500	0.1
A	A	A	$\pm$ (A)
1.002	1.000010	0.001990	0.01
-1.002	-1.000030	-0.00197	0.01
3.002	2.999950	0.002050	0.01
5.003	4.999950	0.003050	0.06
9.003	8.999800	0.003200	0.06

**CERTIFICATE OF CALIBRATION**

CERTIFICATE NUMBER

QC-CAL-25266-20

REQUEST NUMBER: By Mail  
JOB NUMBER: QC-CAL-25266

Page 4 of 5

**Measurement Data for AC Current @ 50 Hz**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu\text{A}$	$\mu\text{A}$
0	0

Before Adjustment	After Adjustment
$\mu\text{A}$	$\mu\text{A}$
0	0

Before Adjustment	After Adjustment
$\mu\text{A}$	$\mu\text{A}$
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
mA	mA	mA	$\pm$ (mA)
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	100.000	0.000	0.1
500	500.030	-0.030	0.1
900	899.900	0.100	0.2
A	A	A	$\pm$ (A)
1.002	1.000100	0.001900	0.01
3.002	3.000040	0.001960	0.01
5.003	5.000040	0.002960	0.06
9.003	9.001000	0.002000	0.06

**Measurement Data for AC Current @ 1 kHz**

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\mu\text{A}$	$\mu\text{A}$
0	0

Before Adjustment	After Adjustment
$\mu\text{A}$	$\mu\text{A}$
0	0

Before Adjustment	After Adjustment
$\mu\text{A}$	$\mu\text{A}$
0	0

**Measurement Data for AC Current @ 1 kHz**

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
mA	mA	mA	$\pm$ (mA)
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100.30	100.000	0.300	0.1
499.40	500.000	-0.600	0.1
599.94	600.000	-0.060	0.2
A	A	A	$\pm$ (A)
1.002	1.000000	0.002000	0.01
3.002	3.000000	0.002000	0.01
5.003	5.000000	0.003000	0.06
9.006	9.000000	0.006000	0.06

**CERTIFICATE OF CALIBRATION**

CERTIFICATE NUMBER



Address : North Rumaila, Al Takamul Yard  
Contact : +9647810009138

**CERTIFICATE OF CALIBRATION**

CERTIFICATE NUMBER

QC-CAL-25266-20

Page 5 of 5

 REQUEST NUMBER: By Mail  
 JOB NUMBER: QC-CAL-25266

**Measurement Data for Resistance**  
 Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
$\Omega$	$\Omega$
0	0

Before Adjustment	After Adjustment
k $\Omega$	k $\Omega$
0	0

Before Adjustment	After Adjustment
M $\Omega$	M $\Omega$
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
$\Omega$	$\Omega$	$\Omega$	$\pm (\Omega)$
1.1	1.00010	0.09990	0.08
10.1	9.99990	0.10010	0.10
100.1	99.9994	0.10060	0.08
599.5	600.0000	-0.50000	0.08
k $\Omega$	k $\Omega$	k $\Omega$	$\pm (k\Omega)$
1.002	1.000000	0.002000	0.01
10.002	9.99996	0.002040	0.01
100.1	100.0006	0.099400	0.06
500.1	499.9993	0.100700	0.05
599.5	600.0000	-0.500000	0.08
M $\Omega$	M $\Omega$	M $\Omega$	$\pm (M\Omega)$
1.001	1.000001	0.000999	0.12
10.000	10.00003	-0.00003	0.01
30.005	30.0000	0.0050	0.02
59.45	60.0000	-0.5500	0.02


 Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

**DEVIATION FROM STANDARD METHOD :** None

**REMARK (S) :**

- The results are as found (no adjustment done).
- The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY	LAB INCHARGE	CLIENT
 Hussein Alaa		 LAB INCHARGE	Asjad Rafiq