

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 15, 2025

Page 1 of 1

REQUEST NUMBER	: By Contract	CUSTOMER DETAILS	
JOB NUMBER	: QC-CAL-25268	Name	Halliburton Worldwide IRAQ (Baroid)
CERTIFICATE NUMBER	: QC-CAL-25268-10	Address	Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
 Manufacturer : UL
 Model : NA
 Serial Number : **BAR-G-59**
 Type : Analog
 Range : 0 to 1500 psi
 Calibrated Range : 0 to 1500 psi
 Resolution : 50 psi
 Accuracy : NA
 Tolerance : ± 1.0 % of Full Scale
 As found : In Tolerance
 Calibration Date : July 15, 2025
Calibration Due : July 14, 2026
 Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure **QC/CP/R/03 Rev. 01** And Halliburton Doc. No **ST-GL-HAL-PSL-900**
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	05 Nov, 2024	04 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.0	0.0	0.0
250	251.6	-1.60	-0.11
500	502.0	-2.00	-0.13
1000	1004.0	-4.00	-0.27
1500	1507.0	-7.00	-0.47

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY	REVIWED & APPROVED BY / LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafi	