Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq
●Phone : +964 7810009138 ● www.qualitycontrol-iraq.com ● E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 15, 2025 **Page 1 of 1**

		CUSTOMER DETAILS	
REQUEST NUMBER	: By Contract	Name	**Halliburton Worldwide IRAQ (Baroid)**
JOB NUMBER	: QC-CAL-25268	Address	Western Burjesia, Oil Street, District Zubair-South Iraq
CERTIFICATE NUMBER	: **QC-CAL-25268-13**		

Description	: **Pressure Gauge**		
Manufacturer	: USG		
Model	: NA		
Serial Number	: **BAR-G-8**		
Type	: Analog		
Range	: 0	to	200 psi
Calibrated Range	: 0	to	200 psi
Resolution	: 5	psi	.
Accuracy	: NA		
Tolerance	± 1.0	% of Full Scale	
As found	: In Tolerance		
Calibration Date	: July 15, 2025		
Calibration Due	: **July 14, 2026**		
Last Calibration Date	: NA		

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure **QC/CP/R/03 Rev. 01** And Halliburton Doc. No **ST-GL-HAL-PSL-900** The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	05 Nov, 2024	04 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.0	0.0	0.0
50	49.7	0.30	0.15
100	99.9	0.10	0.05
150	150.2	-0.20	-0.10
200	200.4	-0.40	-0.20

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

☑ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIWED & APPROVED BY LAB INCHARGE	CLIENT
Hussein Alaa	LAB INCHARGE Asjad Rafi	