

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 15, 2025

Page 1 of 1

REQUEST NUMBER	: By Contract	CUSTOMER DETAILS Name Halliburton Worldwide IRAQ (Baroid) Address Western Burjesia, Oil Street, District Zubair-South Iraq	
JOB NUMBER	: QC-CAL-25268		
CERTIFICATE NUMBER	: QC-CAL-25268-14		

Description : Pressure Gauge
Manufacturer : MC
Model : NA
Serial Number : BAR-G-113
Type : Analog
Range : 0 to 30 psi
Calibrated Range : 0 to 30 psi
Resolution : 1 psi
Accuracy : NA
Tolerance : ± 1.0 % of Full Scale
As found : In Tolerance
Calibration Date : July 15, 2025
Calibration Due : **July 14, 2026**
Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C **Humidity** : 30.0 ± 5 % RH **Atm. Pressure** : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure **QC/CP/R/03 Rev. 01** And Halliburton Doc. No **ST-GL-HAL-PSL-900**
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED



DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	05 Nov, 2024	04 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.0	0.0	0.0
7.5	7.53	-0.03	-0.10
15	15.04	-0.04	-0.13
22.5	22.55	-0.05	-0.17
30	30.07	-0.07	-0.23

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIWED & APPROVED BY LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafi	