Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

●Phone : +964 7810009138 ● www.qualitycontrol-iraq.com ● E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 15, 2025			Page 1 of 1

REQUEST NUMBER	:	By Mail	CUSTOMER DETAILS	
JOB NUMBER	:	QC-CAL-25278	Name	: **Halliburton Worldwide IRAQ** (Baroid)
CERTIFICATE NUMBER	:	**QC-CAL-25278-01**	Address	: Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description	:	**Dial Caliper**
Manufacturer	:	YATO
Model	:	YT-7208
Serial Number	:	GS536121
SAP No	:	NA
Type	:	Analog
Range	:	0 to 150 mm
Calibration Range	:	0 to 150 mm
Readability	:	0.02 mm
As Found	:	In Tolerance
Calibration Date	:	July 15, 2025
Calibration Due	:	**14 July 2026**
Last Calibration	:	March 24, 2024

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature   : $22.0 \pm 2°C$ Humidity : 30.0 ± 5 % RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.14

The reported uncertainty is the expanded uncertainty with k=2 according to the

ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 5, 2024	November 4, 2025

CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
mm	mm	mm	±mm
0.00	0.00	0.00	0.000
25.00	25.00	0.00	0.001
50.00	50.00	0.00	0.001
75.00	75.00	0.00	0.001
100.00	100.00	0.00	0.001
149.98	150.00	0.02	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

■ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
Abdulrahman Loay	LAB INCHARGE Asjad Rafiq	

Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138