Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq
●Phone : +964 7810009138 ● www.qualitycontrol-iraq.com ● E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 27, 2025

REQUEST NUMBER	: by mail	CUSTOMER DETAILS	
JOB NUMBER	: QC-CAL-25292	Name	**Halliburton Worldwide IRAQ (Baroid)**
CERTIFICATE NUMBER	: **QC-CAL-25292-06**	Address	Western Burjesia, Oil Street, District Zubair-South Iraq

Description	: **Pressure Gauge**		
Manufacturer	: ASHCROFT		
Model	: NA		
Serial Number	: **BAR-G-51**		
Type	: Analog		
Range	: 0	to	5000 psi
Calibrated Range	: 0	to	5000 psi
Resolution	: 50	psi	.
Accuracy	: NA		
Tolerance	± 1.0	% of Full Scale	
As found	: In Tolerance		
Calibration Date	: July 27, 2025		
Calibration Due	: **July 26, 2026**		

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 20.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	5 Nov, 2024	4 Nov, 2025

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
1250	1258.0	-8.00	-0.2
2500	2513.0	-13.00	-0.3
3750	3765.0	-15.00	-0.3
5000	5020.0	-20.00	-0.4

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

☑ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIWED & APPROVED BY LAB INCHARGE	CLIENT
Abdulrahman Loay	LAB INCHARGE Asjad Rafiq	