

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 5, 2025

| | | |
|--------------------|-------------------|-------------|
| REQUEST NUMBER | : By Contract | APPROVED BY |
| JOB NUMBER | : QC-CAL-25351 | Asjad |
| CERTIFICATE NUMBER | : QC-CAL-25351-24 | |



CUSTOMER DETAILS

Name : **slb Basra**
 Address : North Rumaila Base south Iraq
 Department : Wireline

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
 Manufacturer : NOV Elmar
 Model : EN837-1
 Serial Number : 211252107
 Type : Analog
 Location : Skid-41
 Range : 0 to 6000 psi
 Calibrated Range : 0 to 6000 psi
 Resolution : 100 psi
 Accuracy Class : 1.6
 Tolerance : ±1.6 % of Full Scale
 As found : In Tolerance
 Calibrated By : Hussein Alaa
 Calibration Date : September 5, 2025
Calibration Due : 04 September 2026



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

| DESCRIPTION | SERIAL NO. | MAKE / MODEL | CALIBRATION DATE | CALIBRATION DUE DATE |
|------------------------|------------|---------------|------------------|----------------------|
| Digital Pressure Gauge | 468802 | AMETEK / XP2i | 05 Nov, 2024 | 04 Nov, 2025 |

CALIBRATION TEST RESULTS

| Measured Values on Unit Under Calibration (M) | Ref. Standard Values (A) | Deviation (D=M-A) | Error Percentage |
|---|--------------------------|-------------------|------------------|
| psi | psi | psi | ± |
| 0 | 0.00 | 0.00 | 0.0 |
| 1500 | 1508.00 | -8.00 | -0.1 |
| 3000 | 3011.00 | -11.00 | -0.2 |
| 4500 | 4513.00 | -13.00 | -0.2 |
| 6000 | 6019.00 | -19.00 | -0.3 |

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.



Corporate Partner

Address : North Rumaila, Al Takamul Yard
 Contact : +9647810009138