



## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** September 8, 2025

<b>REQUEST NUMBER</b> : By Contract <b>JOB NUMBER</b> : QC-CAL-25351 <b>CERTIFICATE NUMBER</b> : QC-CAL-25351-32	APPROVED BY:  
--	---

### CUSTOMER DETAILS

**Name** : slb Basra  
**Address** : North Rumaila Base south Iraq  
**Department** : Wireline

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

**Description** : **Pressure Gauge**  
**Manufacturer** : NA  
**Model** : NA  
**Serial Number** : CAL-PG-01  
**Type** : Analog  
**Location** : Skid-14  
**Range** : 0 to 200 psi  
**Calibrated Range** : 0 to 200 psi  
**Resolution** : 10 psi  
**Accuracy Class** : 1.6  
**Tolerance** : ±1.6 % of Full Scale  
**As found** : In Tolerance  
**Calibrated By** : Hussein Alaa  
**Calibration Date** : September 8, 2025  
**Calibration Due** : **07 September 2026**



**DEVIATION FROM STANDARD METHOD** : None

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

**Ambient Temperature** : 25.0 ± 2°C    **Humidity** : 30.0 ± 5 % RH    **Atm. Pressure** : 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01  
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	05 Nov, 2024	04 Nov, 2025

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±
0	0.00	0.00	0.0
50	50.30	-0.30	-0.1
100	100.60	-0.60	-0.3
150	150.80	-0.80	-0.4
200	201.10	-1.10	-0.5

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.

