



## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** September 8, 2025

|   |   |
|---|---|
| <b>REQUEST NUMBER</b> : By Contract         | APPROVED BY  |
| <b>JOB NUMBER</b> : QC-CAL-25351            |   |
| <b>CERTIFICATE NUMBER</b> : QC-CAL-25351-36 |   |



Page 1 of 1

### CUSTOMER DETAILS

**Name** : slb Basra  
**Address** : North Rumaila Base south Iraq  
**Department** : Wireline

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

**Description** : Pressure Gauge  
**Manufacturer** : NOV Elmar  
**Model** : EN837-1  
**Serial Number** : 211752474  
**Type** : Analog  
**Location** : Skid-14  
**Range** : 0 to 10000 psi  
**Calibrated Range** : 0 to 10000 psi  
**Resolution** : 200 psi  
**Accuracy Class** : 2.5  
**Tolerance** : ±2.5 % of Full Scale  
**As found** : In Tolerance  
**Calibrated By** : Hussein Alaa  
**Calibration Date** : September 8, 2025  
**Calibration Due** : 07 September 2026



**DEVIATION FROM STANDARD METHOD** : None

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

**Ambient Temperature** : 25.0 ± 2°C    **Humidity** : 30.0 ± 5 % RH    **Atm. Pressure** : 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01  
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

| DESCRIPTION            | SERIAL NO. | MAKE / MODEL  | CALIBRATION DATE | CALIBRATION DUE DATE |
|------------------------|------------|---------------|------------------|----------------------|
| Digital Pressure Gauge | 468802     | AMETEK / XP2i | 05 Nov, 2024     | 04 Nov, 2025         |

### CALIBRATION TEST RESULTS

| Measured Values on Unit Under Calibration (M) | Ref. Standard Values (A) | Deviation (D=M-A) | Error Percentage |
|---|--------------------------|-------------------|------------------|
| psi   | psi                      | psi               | ±                |
| 0   | 0.00                     | 0.00              | 0.0              |
| 2400  | 2416.00                  | -16.00            | -0.2             |
| 5000  | 5024.00                  | -24.00            | -0.2             |
| 7400  | 7427.00                  | -27.00            | -0.3             |
| 10000   | 10039.00                 | -39.00            | -0.4             |

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.

