

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: January 11, 2026

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY LAB IN CHARGE QC
JOB NUMBER : QC-CAL-26019	
CERTIFICATE NUMBER : QC-CAL-26019-04	

CUSTOMER DETAILS

Name : **Daqing Drilling Company Iraq**
 Address : Basra, South Iraq
 Department : DQ038

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **RCD Tester**
 Manufacturer : UNI-T
 Model : UT-582
 Serial Number : **C230155897**
 Type : Digital
 Range : 10 to 500 mA
 Calibrated Range : 10 to 500 mA
 Readability : 0.01 mA
 Calibration Date : January 11, 2026
Calibration Due : January 10, 2027
 Last Calibration : NA



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 20.0 ± 1°C Humidity : 35.0 ± 5 % RH Atm. Pressure : 1013.6 ± 10 hPa

CALIBRATION METHOD

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Electrical Test Equipment Calibrator	TRANSMILLE	3200A	2806902	November 3, 2025	November 2, 2026

CALIBRATION TEST RESULTS

Range Function	Selected Values (A)	Measured Values (M)	Deviation (D=M-A)	Expanded Uncertainty (U)	
	mA	mA	mA	±	mA
x1/2	10.00	10.02	0.02	0.05	
x1/2	20.0	19.98	-0.02	0.05	
x1	100.00	99.96	-0.04	0.05	
x1	300.0	300.04	0.04	0.05	
x5	20	20.02	0.02	0.05	
x5	30	30.03	0.03	0.05	

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.