

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 22, 2026

Page 1 of 1

REQUEST NUMBER : by mail
JOB NUMBER : QC-CAL-26154
CERTIFICATE NUMBER : QC-CAL-26154-17

APPROVED BY LAB INCHARGE QC



CUSTOMER DETAILS

Name : Egyptian Maintenance Company(EMC) IRAQ
Address : Iraq, Basra

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Adhesion Tester
Manufacturer : Elcometer
Model : F510-20S
Serial Number : BC04849
Type : Analog
Range : 300 to 3600 psi
Tolerance : $\pm 1.5\%$ (as Per Manufacturer)
As found : Good
Calibration Date : September 22, 2026
Calibration Due : September 21, 2027



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : $40.0 \pm 2^{\circ}\text{C}$ Humidity : $25.0 \pm 5\% \text{ RH}$ Atm. Pressure : $1012.0 \pm 10 \text{ hPa}$

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/06 Rev. 01

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Push off Adhesion Tester	NL58390	Elcometer / 508	November 3, 2025	November 2, 2026

CALIBRATION TEST RESULTS

the Unit Under Calibration and the Reference Equipment, all use same surfaces During Do the Test and at the Same Temperature.

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit ($\pm 1.5\%$)
psi	psi	psi	%
850.0	844.0	6.0	1.06
1250.0	1259.0	-9.0	-1.08

Results:

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.