

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 22, 2026

Loading Note No	: by mail	APPROVED BY: L. AL-MUHAMMADI QC
JOB NUMBER	: QC-CAL-26154	
CERTIFICATE NUMBER	: QC-CAL-26154-24	

CUSTOMER DETAILS

Name : Egyptian Maintenance Company(EMC) IRAQ
Address : Iraq, Basra, North Rumaila

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Electrode Oven
Manufacturer : Phoenix
Model : 300-High Temp
Serial Number : 0526513366
Type : Analog
Range : (122- 662)°F
As found : In Tolerance
Calibrated By : Hussein Alaa
Calibration Date : September 22, 2026
Calibration Due : September 21, 2027



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 40.0 ± 1°C Humidity : 25.0 ± 5 % RH Atm. Pressure 1013.0 ± 10 hPa

CALIBRATION METHOD

This calibration was carried out in compliance with ISO/IEC 17025, 2017 & ISO 9001, 2015. Calibration of above mentioned instrument has performed against reference standard digital Thermometer with Type K Thermocouple. The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	DUE DATE
Digital Thermometer with K-Type Thermocouple	PATOS	DE-305	G0009036	3-Nov-25	2-Nov-26

CALIBRATION TEST RESULTS

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
°C	°C	°C	± (°C)
122.0	122.25	-0.25	2.0
240.0	240.66	-0.66	2.0
360.0	361.00	-1.00	2.0
580.0	581.32	-1.32	2.0
662.0	663.50	-1.50	2.0

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) : ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.