

Calibration Certificate

Al Takamul Yard, North Rumailah Iraq

● Phone : +964 7810009138 ● www.qualitycontrol-iraq ● E-mail: op@qualitycontrol-iraq

Date of Issue: January 3, 2024

Page 1 of 1

REQUEST NUMBER : By Mail QC JOB NUMBER : QC/JN/24/00101 CERTIFICATE NUMBER : QC240103-01	CUSTOMERS DETAILS Name : Halliburton Worldwide-Iraq Branch (Sperry) Address : Oil Street, Western Burjessia Basra South Iraq
--	--

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Torque Wrench Type of Indication : Analog Manufacturer : LICOTA Model : AQL-N3110V Serial Number : 17065620 SAP No. : 300314707 Calibrated Range : 20 to 110 N.m Resolution : 0.5 N.m Tolerance : ± 4 % As Found : Good	Calibration Date : January 3, 2024 Calibration Due : January 2, 2025 Last Calibration : NEW
---	---



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C ± 2°C Relative Humidity : 45 %RH ± 5% RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with International Calibration standard # ISO 6789-1:2017

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE / MODEL	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Torque Wrench Calibrator	Nobar / Pro Test 1500	90962	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measurement Data For Torque

Clockwise Measurements

Readings on UUC	Readings on Ref. Standard (Avg. of 5 Measurements)	Deviation	Error Percentage (Tolerance ±4%)	Uncertainty (95 % C.L)
N.m	N.m	N.m	%	± N.m
20.0	20.30	-0.30	-1.5	0.12
65.0	66.30	-1.30	-2.0	0.12
110.0	112.50	-2.50	-2.3	0.12

Counter Clockwise Measurements

20.0	20.40	-0.40	-2.0	0.12
65.0	66.80	-1.80	-2.8	0.12
110.0	112.70	-2.70	-2.5	0.12

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) : The results are as found (no adjustment done).
 The results are post adjustment.

CALIBRATED BY  Mahdi Halim	REVISED & APPROVED BY LAB INCHARGE  LAB INCHARGE	CLIENT Asjad Rafiq
---	---	---------------------------