

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: February 19, 2024

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY
JOB NUMBER	: QC/JN/24/00145	
CERTIFICATE NUMBER	: QC240219-10	



CUSTOMER DETAILS

Name : **slb Basra**
 Address : North Rumaila Base south Iraq
 Department : Testing

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
 Manufacturer : CP
 Model : 680
 Serial Number : **21021070006**
 Location : Hydraulic Pressure Pump
 Type : Digital
 Range : 0 to 10000 psi
 Calibrated Range : 0 to 10000 psi
 Resolution : 1 psi
 Tolerance : 0.25 % from Each Reading
 As found Condition : Good
 Calibrated By : Asjad Rafiq
 Calibration Date : February 19, 2024
Calibration Due : 18 February 2025
 Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
psi	psi	psi	± psi
0	0.00	0.00	0.00
2500	2503.00	-3.00	6.25
5000	5005.00	-5.00	12.50
7500	7509.00	-9.00	18.75
10000	10013.00	-13.00	25.00

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

