

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: March 21, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY: _____
JOB NUMBER : QC/JN/24/00161	
CERTIFICATE NUMBER : QC240321-02	



CUSTOMER DETAILS

Name : **slb Basra**
 Address : North Rumaila Base south Iraq
 Department : Testing

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
 Manufacturer : BADOTHERM
 Model : AISI 316
 Serial Number : **V1544-1752**
 Location : ESD CB-793
 Type : Analog
 Range : 0 to 400 bar
 Calibrated Range : 0 to 400 bar
 Resolution : 5 bar
 Tolerance : 1 % from Each Reading
 As found Condition : Good
 Calibrated By : Asjad Rafiq
 Calibration Date : March 21, 2024
Calibration Due : 20 March 2025
 Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01
 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
bar	bar	bar	± bar
0	0.00	0.00	0.00
100	100.68	-0.68	1.00
200	200.84	-0.84	2.00
300	300.97	-0.97	3.00
400	401.13	-1.13	4.00

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

