

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: March 26, 2024

Page 1 of 1

REQUEST NUMBER	: By Contract	CUSTOMER DETAILS	
JOB NUMBER	: QC/JN/24/00167	Name	: Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER	: QC240326-01	Address	: Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Relief Valve**
 Manufacturer : VEEKAY PROCESS INSTRUMENT
 Model : CS
 Serial Number : V-5001
 SAP : 300231267
 Location : **Flow Loop Test Equipment: LEPU 6666 022**
 Set/Pressure : 10.5 kg/cm2
 Tolerance : ± 3%
 As Found : Good
 Calibration Date : March 26, 2024
Calibration Due : **25 March 2025**
 Last Calibration : March 6, 2023



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

Standard Method API 527 was used to test the unit under test. Test pressure was applied on inlet of unit under test and test was repeated 3 times. Identical repeatings were carried out as recommended in used standard.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	1505119397	Sino Instruments / HS108	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Set Pressure on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
kg/cm2	kg/cm2	kg/cm2	± kg/cm2
10.5	10.7	-0.2	0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- The results are as found (no adjustment done).
- The results are post adjustment.

CALIBRATED BY  Mahdi Halim	REVIEWED & APPROVED BY  LAB INCHARGE Asjad Rafiq	CLIENT
---	--	--------

