

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: March 26, 2024

Page 1 of 1

REQUEST NUMBER : By Contract		CUSTOMER DETAILS	
JOB NUMBER : QC/JN/24/00167		Name	Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER : QC240326-03		Address	Western Burjesia, Oil Street, District Zubair-South Iraq

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**  
 Manufacturer : Baumer  
 Model : SS 316L  
 Serial Number : **L.083.59-00949**  
 SAP : **300231665**  
 Type : Analog  
 Location : **Flow Loop Test Equipment: LEPU 6666 022**  
 Range : 0 to 15 psi  
 Calibrated Range : 0 to 15 psi  
 Resolution : 0.2 psi  
 Tolerance : ± 1.6 % from Each Reading  
 As found : Good  
 Calibration Date : March 26, 2024  
**Calibration Due : 25 March 2025**  
 Last Calibration Date : March 6, 2023



**DEVIATION FROM STANDARD METHOD : None**

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED




DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
psi	psi	psi	± psi
0	0.00	0.00	0.00
4	4.00	0.00	0.06
8	8.00	0.00	0.13
12	12.01	-0.01	0.19
15	15.01	-0.01	0.24

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY	LAB INCHARGE	CLIENT
 Mahdi Halim	 Asjad Rafiq	 LAB INCHARGE	