

## Certificate of Calibration

AlTakamul Yard, North Rumaila-South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** April 24, 2024

Page 1 of 1

REQUEST NUMBER	By Mail	APPROVED BY: [Signature] CALIBRATION ENGINE-QC
JOB NUMBER	QC/JN/24/00189	
CERTIFICATE NUMBER	QC240424-11	



### CUSTOMER DETAILS

Name : **Daqing Drilling Company IRAQ**  
 Address : Basra, South Iraq  
 Rig : DQ037

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Portable H2S Gas Detector**  
 Manufacturer : HONEYWELL  
 Model : BW Clip  
 Serial Number : 5220BWC01230213689  
 Type : Digital  
 Calibrated Range : 25 PPM H2S  
 As found : Good  
 Calibrated By : Asjad Rafiq  
 Calibration Date : April 24, 2024

Alarm Details	Low	High
H2S PPM	10	15



**Calibration Due**

July 23, 2024

3 Months Validity from the date of Calibration

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C ± 2 °C  
 Relative Humidity : 30 %RH ± 5 %RH

### CALIBRATION METHOD

A reference cylinder consisting of Sulphur Dioxide (H2S) Gas is used , which are detected by gas detector by diffusion.

### TRACEABILITY

The measurements made by QC Calibration & Testing Labs are traceable to NIST, vide Calgaz Certificate of Analysis  
 The measurements made by QC Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

### REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE	GAS EXPIRY DATE
Multigas Calibration cylinder (25 PPM H2S)	950-106554	CALGAZ	21/08/2025

### CALIBRATION TEST RESULTS

Gas Name	Values of UUC	Values on Reference Standard	Error in Reading of UUC	Uncertainty of Measurement
H2S Gas (25 ppm)	26.0	25	-1	± 0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).  
 The results are post adjustment.