

Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue : April 27, 2024

REQUEST NUMBER : By Contract

APPROVED BY : [Signature]

JOB NUMBER : QC/JN/24/00193

CERTIFICATE NUMBER : QC240427-56

**CUSTOMER DETAILS**

Name : DQ (MUD UNIT)

Address : Rumaila Base south Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Labortary Stirrer

Manufacturer : CORNING

Model : PC4100

Serial Number : 155214013076

Range : 60- 1150 RPM

Resolution : 10 RPM

As Found : In Tolerance

Calibrated By : Asjad Rafiq 60

Calibration Date : April 27, 2024

Calibration Due : April 26, 2025

Last Calibration : NA

**ENVIRONMENTAL CONDITIONS DURING TEST**

Ambient Temperature : 22 °C ± 2 °C

Relative Humidity : 40 %RH ± 5 %RH

CALIBRATION METHOD

The above equipment has been calibrated by measuring RPM against a Reference Standard Non Contact Tachometer

The deviations of the measurement obtained from UUC with respect to reference standards are determined to obtain the error .

TRACEABILITY

The measurements made by Issued by: Quality Control Calibration & Testing Lab., realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Digital Tachometer	8598564223	FLUKE / 930	November 20, 2023	November 19, 2024

CALIBRATION TEST RESULTS

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95% C.L)
RPM	RPM	RPM	± RPM
60	61.2	1.2	2.0
200	201.0	1.0	2.0
400	402.0	2.0	2.0
600	602.0	2.0	2.0
900	904.0	4.0	2.0
1150	1154.0	4.0	2.0

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has PASSED its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

REMARK (S) :
☒ The results are as found (no adjustment done).
☐ The results are post adjustment.Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138