

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: May 2, 2024

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY
JOB NUMBER	: QC/JN/24/00197	
CERTIFICATE NUMBER	: QC240502-04	



CUSTOMER DETAILS

Name : **slb Basra**
 Address : North Rumaila Base south Iraq
 Department : Wireline

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
 Manufacturer : **NOV**
 Model : EN837-1
 Serial Number : **172952536**
 Location : PCE # 32
 Type : Analog
 Range : 0 to 6000 psi
 Calibrated Range : 0 to 6000 psi
 Resolution : 100 psi
 Accuracy : 1.6
 Tolerance : ±1.6 % of Full Scale
 As found : In Tolerance
 Calibrated By : Asjad Rafiq
 Calibration Date : May 2, 2024
Calibration Due : **01 May 2025**
 Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
1500	1513.00	-13.00	-0.2
3000	3018.00	-18.00	-0.3
4500	4522.00	-22.00	-0.4
6000	6025.00	-25.00	-0.4

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

