

# **Certificate of Calibration**

AlTakamul Yard, North Rumaila-South Iraq •Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com Date of Issue: May 11, 2024 Page 1 of 1 REQUEST NUMBER By Mail APPROVED GE-OC QC/JN/24/00209 JOB NUMBER QC240511-15 ERTIFICATE NUMBER **CUSTOMER DETAILS** : Daqing Drilling Company IRAQ Name Address : Basra, South Iraq : DQ011 Rig EQUIPMENT IDENTIFICATION AND SPECIFICATIONS **Portable H2S Gas Detector** Description Manufacturer HONEYWELL Model **BW** Clip 5220BWC01232801062 Serial Number Type Digital 25 PPM H2S Alarm Details Low Calibrated Range High As found H2S PPM 10 15 Good Calibrated By Asjad Rafiq May 11, 2024 Calibration Date August 10, 2024 **Calibration Due** 3 Months Validity from the date of Calibration

#### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature	:	22 °C	±	2 °C
Relative Humidity		30 %RH	±	5 %RH

#### **CALIBRATION METHOD**

A reference cylinder consisting of Sulpher Dioxide (H2S) Gas is used , which are detected by gas detector by diffusion.

#### TRACEABILITY

The measurements made by QC Calibration & Testing Labs are traceable to NIST, vide Calgaz Certificate of Analysis The measurements made by QC Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

### **REFERENCE EQUIPMENT USED :**

DESCRIPTION	SERIAL NO.	MAKE	GAS EXPIRY DATE
Multigas Calibration cylinder (25 PPM H2S)	950-106554	CALGAZ	21/08/2025

## CALIBRATION TEST RESULTS

Gas Name	Values of UUC	Values on Reference Standard	Error in Reading of UUC	Uncertainty of Measurement
H2S Gas (25 ppm)	25.0	25	0	$\pm 0.2$

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☑ The results are as found (no adjustment done).
- □ The results are post adjustment.



180

9001