

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: May 12, 2024

Page 1 of 1

REQUEST NUMBER	: By Contract	CUSTOMER DETAILS	
JOB NUMBER	: QC/JN/24/00210	Name	Halliburton Worldwide IRAQ (WPS)
CERTIFICATE NUMBER	: QC240512-02	Address	Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
Manufacturer : NOV ASEP ELMAR
Model : EN 837-1
Serial Number : **125152189**
Type : Analog
Location : PCE 9-4 (SAP # 12253877)
Range : 0 to 15000 psi
Calibrated Range : 0 to 15000 psi
Resolution : 100 psi
Accuracy : 1.0
Tolerance : ± 1.0 % of Full Scale
As found : In Tolerance
Calibration Date : May 12, 2024
Calibration Due : **11 May 2025**
Last Calibration Date : August 30, 2023



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 35.0 ± 2°C Humidity : 30.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
0	0	0	%
0	0.00	0.00	0.0
3700	3695.00	5.00	0.0
7500	7498.00	2.00	0.0
11200	11206.00	-6.00	0.0
15000	15010.00	-10.00	-0.1

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED LAB INCHARGE	CLIENT
 Mahdi Halim	 LAB INCHARGE Asjad Rafiq	ALI Talib HB48903 Date: 04/04/2024 Signature  Halliburton