

Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: May 16, 2024

Page 1 of 2

REQUEST NUMBER	: By Email	APPROVED BY: LAB IN CHARGE QC
JOB NUMBER	: QC/JN/24/00216	
CERTIFICATE NUMBER	: QC240516-01	



CUSTOMER DETAILS

Name : slb Basra
Address : North Rumaila Base south Iraq
Department : Coil Tubing

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Load Cell
Manufacturer : NA
Model : NA
Asset No. : NA
Serial Number : 675495



Calibrated Range:

Force : 0 K lbs to 40 K lbs

As Found : In Tolerance

Location : QC Labs North Rumailah

Calibrated By : Asjad Rafiq

Calibration Date : May 16, 2024

Calibration Due : Recommended Validity : one (1) year from the date of calibration, (Where Required) May 15, 2025

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C ± 2 °C
Relative Humidity : 45 %RH ± 5 %RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with International Calibration Procedure # ASTM E74-13a

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

TRACEABILITY

The measurements made by Quality Control Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.



Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138

CERTIFICATE OF CALIBRATIONREQUEST NUMBER : By Email
JOB NUMBER : QC/JN/24/00216

CERTIFICATE NUMBER

QC240516-01

Page 2 of 2

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Load Cell	ZEMIC	BM14A-C3-40T	TA906164	November 2, 2023	November 1, 2024

CALIBRATION TEST RESULTS

Measurement Data For Force

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
K lbs	K lbs
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L)
K lbs	K lbs	K lbs	±K lbs
0	0.00	0.00	0
10	10.04	-0.04	0.2
20	20.06	-0.06	0.2
30	30.05	-0.05	0.2
40	40.07	-0.07	0.2

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

~ END OF DOCUMENT ~

This certificate is issued in accordance with the laboratory accreditation requirements as per ISO/IEC17025:2017.

This certificate may not be reproduced other than in full, unless permission for the publication of an approved extract has been obtained in writing from the director of QC. It does not of itself impute to the subject of the calibration any attributes beyond those shown by the data contained herein.