

Certificate of Calibration


AlTakamul Yard, North Rumaila-South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue:

May 26, 2024

Page 1 of 1

REQUEST NUMBER	By Mail	 APPROVED FOR CALIBRATION CALGAS TEST 2 For Engineering Services
JOB NUMBER	QC/JN/24/00223	
CERTIFICATE NUMBER	QC240526-07	

CUSTOMER DETAILS

Name : Zhongman Petroleum Engineering
Address : Basra, South Iraq
Rig : ZPEC025

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Portable H2S Gas Sensor
Manufacturer : HONEYWELL
Model : BW Clip
Serial Number : 5220BWC01221316015
Type : Digital
Calibrated Range : 25 PPM H2S
As found : Good
Calibrated By : Asjad Rafiq
Calibration Date : May 26, 2024

Alarm Details	Low	High
H2S PPM	10	15

**Calibration Due**

August 25, 2024

3 Months Validity from the date of Calibration

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 °C ± 2 °C
Relative Humidity : 30 %RH ± 5 %RH

CALIBRATION METHOD

A reference cylinder consisting of Sulphur Dioxide (H2S) Gas is used , which are detected by gas detector by diffusion.

TRACEABILITY

The measurements made by QC Calibration & Testing Labs are traceable to NIST, vide Calgaz Certificate of Analysis
The measurements made by QC Calibration & Testing Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE	GAS EXPIRY DATE
Multigas Calibration cylinder (25 PPM H2S)	950-106554	CALGAZ	21/08/2025

CALIBRATION TEST RESULTS

Gas Name	Values of UUC	Values on Reference Standard	Error in Reading of UUC	Uncertainty of Measurement
H2S Gas (25 ppm)	25.0	25	0	± 0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.