

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

•Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: May 27, 2024				Page 1 of 1		
REQUEST NUMBER		: By Contrac	ct	APPROVED BY		
JOB NUMBER		: QC/JN/24/	00223	CAL JEL		
CERTIFICATE NUMBE	ER	: QC240527	/-04			
				AL CONTROL		
			CUSTOMER DET	AILS		
Name	: Zhong	man Petroleum I	Engineering			
Address	: Basra,	South Iraq		chgineeting		
Department	: ZPEC()25		<u> </u>		
		EQUIPM	IENT IDENTIFICATION A	ND SPECIFICATIONS		
Description	: Pressu	re Gauge				
Manufacturer	: MC			見た物感見		
Model	: 200032	28				
Serial Number	: 161057	705				
Location	: Koom	ey Unit				
Туре	: Analog	5		画編編編版		
Range	: 0	to	360 psi			
Calibrated Range	: 0	to	360 psi			
Resolution	: 5	psi		DEVIATION FROM STANDARD METHOD : None		
Accuracy Class	: 1.6					
Tolerance	: 1	% of Full S	cale	REMARK (S) :		
As found	: In Tole	erance		1. No adjustments were made prior to calibration.		
Calibrated By	: Asjad l	Rafiq		2. Overload test was not performed.		
Calibration Date	: May 2'	7, 2024		3. No accessories were fitted.		
Calibration Due	: 26 May 2025			4. Values were rounded in computation.		
Last Calibration Date	: NA					

ENVIRONMENTAL CONDITIONS DURING TEST						
Ambient Temperature	:	25.0 ± 2°C	Humidity :	$44.0~\pm 5~\%~RH$	Atm. Pressure	1011.0 ± 10 hPa
CALIBRATION METHOD						

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01 The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard

having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED					
DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE	
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024	

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±
0	0	0.0	0.0
50	50.1	-0.1	0.0
100	100.2	-0.2	-0.1
200	200.5	-0.5	-0.1
300	301.1	-1.1	-0.3
360	361.4	-1.4	-0.4


Corporate

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has PASSED its Calibration.

- - $\stackrel{[]}{\frown}$ The results are as found (no adjustment done).
 - The results are post adjustment.











OHSAS 18001





hoc .



Address : North Rumaila, Al Takamul Yard Contact : +9647810009138