

## Certificate of Calibration


Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** May 27, 2024

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY	LABORATORY CONTROL
JOB NUMBER	: QC/JN/24/00223		
CERTIFICATE NUMBER	: QC240527-07		



### CUSTOMER DETAILS

**Name** : Zhongman Petroleum Engineering  
**Address** : Basra, South Iraq  
**Department** : ZPEC025

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

**Description** : Pressure Gauge  
**Manufacturer** : MC  
**Model** : XRT  
**Serial Number** : 17051238  
**Location** : CATWALK  
**Type** : Analog  
**Range** : 0 to 40 MPa  
**Calibrated Range** : 0 to 40 MPa  
**Resolution** : 1 MPa  
**Accuracy Class** : 2.5  
**Tolerance** : 1 % of Full Scale  
**As found** : In Tolerance  
**Calibrated By** : Asjad Rafiq  
**Calibration Date** : May 27, 2024  
**Calibration Due** : 26 May 2025  
**Last Calibration Date** : NA



**DEVIATION FROM STANDARD METHOD** : None

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

**Ambient Temperature** : 25.0 ± 2°C    **Humidity** : 44.0 ± 5 % RH    **Atm. Pressure** : 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01  
The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
MPa	MPa	MPa	±
0	0	0.00	0.0
5	4.93	0.07	0.2
10	9.97	0.03	0.1
20	20.05	-0.05	-0.1
30	30.10	-0.10	-0.3
40	40.09	-0.09	-0.2

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).  
☐ The results are post adjustment.