

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** May 27, 2024

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY: 
JOB NUMBER	: QC/JN/24/00223	
CERTIFICATE NUMBER	: <b>QC240527-28</b>	

### CUSTOMER DETAILS

Name : **Zhongman Petroleum Engineering**  
Address : Basra, South Iraq  
Department : ZPEC025

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**  
Manufacturer : NA  
Model : NA  
Serial Number : **CAL24016**  
Location : **ACETYLENE CYLINDER**  
Type : Analog  
Range : 0 to 35 psi  
Calibrated Range : 0 to 35 psi  
Resolution : 1 psi  
Accuracy Class : 1  
Tolerance : 1 % of Full Scale  
As found : In Tolerance  
Calibrated By : Asjad Rafiq  
Calibration Date : May 27, 2024  
**Calibration Due** : **26 May 2025**  
Last Calibration Date : NA



**DEVIATION FROM STANDARD METHOD** : None

**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

### CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

### CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±
0	0	0.00	0.0
5	5.1	0.0	-0.1
10	10.1	-0.1	-0.2
20	20.1	-0.1	-0.3
30	30.1	-0.1	-0.3
35	35.2	-0.1	-0.4

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).  
☐ The results are post adjustment.



Address : North Rumaila, Al Takamul Yard  
Contact : +9647810009138

