


Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: May 27, 2024

Page 1 of 1

REQUEST NUMBER	: By Contract	APPROVED BY: 
JOB NUMBER	: QC/JN/24/00223	
CERTIFICATE NUMBER	: QC240527-32	

CUSTOMER DETAILS

Name : **Zhongman Petroleum Engineering**
Address : Basra, South Iraq
Department : ZPEC025

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
Manufacturer : WIKA
Model : NA
Serial Number : **CAL24018**
Location : **SPARE**
Type : Analog
Range : 0 to 230 psi
Calibrated Range : 0 to 230 psi
Resolution : 5 psi
Accuracy Class : 1
Tolerance : 1 % of Full Scale
As found : In Tolerance
Calibrated By : Asjad Rafiq
Calibration Date : May 27, 2024
Calibration Due : **26 May 2025**
Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure : 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	±
0	0	0.00	0.0
50	50.2	-0.2	-0.1
100	100.6	-0.6	-0.2
150	150.7	-0.7	-0.3
200	200.7	-0.7	-0.3
230	230.8	-0.8	-0.4

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.



Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138

