

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 16, 2024

Page 1 of 1

REQUEST NUMBER	:	By Mail	CUSTOMER DETAILS	
JOB NUMBER	:	QC/JN/24/00282	Name	: Halliburton Worldwide IRAQ (Sperry)
CERTIFICATE NUMBER	:	QC240716-02	Address	: Western Burjesia, Oil Street, District Zubair-South Iraq

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Digital Vernier Caliper**
 Manufacturer : EGA Master
 Model : NA
 Serial Number : 005141
 SAP No : NA
 Type : Digital
 Range : 0 to 6 inch
 Calibration Range : 0 to 6 inch
 Readability : 0.0005 inch
 As Found : Good
 Calibration Date : July 16, 2024
Calibration Due : 15 July 2025
 Last Calibration : August 1, 2023



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 40.0 ± 5 % RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME B89.1.14
 The reported uncertainty is the expanded uncertainty with k=2 according to the
 ISO/IEC "Guide to the expansion of uncertainty in measurement" providing a level of confidence of approximately 95%.

REFERENCE EQUIPMENT USED :


DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	DUE DATE
Gauge Block Set	87300	Mitutoyo / 99MAK004B2	November 15, 2023	November 14, 2024

CALIBRATION TEST RESULTS

Ref. Standard Values (A)	Measured Values on Unit Under Calibration (M)	Deviation (D=M-A)	Expanded Uncertainty (U)
inch	inch	inch	±inch
0.0000	0.0000	0.0000	0.000
1.0000	1.0000	0.0000	0.001
2.0000	2.0000	0.0000	0.001
4.0000	4.0000	0.0000	0.001
6.0000	6.0000	0.0000	0.001

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	



Address : North Rumaila, Al Takamul Yard
 Contact : +9647810009138