

Certificate of Calibration

Al Takamul Yard, North Rumailah, South IRAQ

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: June 24, 2024

Page 1 of 1

REQUEST NUMBER	:	By Mail	APPROVED	ENGINE QC
JOB NUMBER	:	QC/JN/24/00253		
CERTIFICATE NUMBER	:	QC240624-01		



CUSTOMER DETAILS

Name : Egyptian Maintenance Company(EMC) IRAQ
Address : Iraq, Basra, North Rumaila

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : AC DC YOKE
Manufacturer : PARKER
Model : DA 400S
Serial Number: : 29352
Calibrated By : Asajd Rafiq
Calibration Date : June 24, 2024
Calibration Due : **23 December 2024**
Calibration Site : QC Labs North Rumaila



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 42.0 ± 5 % RH Atm. Pressure ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with ASME BPVC-V

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	Calibration Date	Due Date
Load Test Plate	APT 1515	15 Dec, 2023	14 Dec, 2024

CALIBRATION TEST RESULTS

Calibrated Values of Rectangular Plate	PROCEDURE APLIED	Results
lbs		Pass/Fail
10.00	Yoke Energized, Plate lifted and held firmly.	Pass

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- 1 No adjustments were made prior to calibration.
- 2 Overload test was not performed.
- 3 No accessories were fitted.
- 4 Values were rounded in computation.



Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138



Al Takamul Company For Engineering Services

Quality Control – Iraq

This certificate is issued in accordance with the laboratory accreditation requirements as per ISO/IEC17025:2017. This certificate may not be reproduced other than in full, unless permission for the publication of an approved extract has been obtained in writing from the director of QC. It does not of itself impute to the subject of the calibration any attributes beyond those shown by the data contained herein.



Corporate
Partner

Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138