


## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

**Date of Issue:** July 4, 2024

Page 1 of 1

Loading Note No	: By Contract	APPROVED BY 
JOB NUMBER	: QC/JN/24/00265	LABORATORY USE QC
CERTIFICATE NUMBER	: <b>QC240704-92</b>	



**CUSTOMER DETAILS**

Name : **Zhongman Petroleum Engineering**  
 Address : Basra, South Iraq  
 Department : ZPEC023

**EQUIPMENT IDENTIFICATION AND SPECIFICATIONS**

Description : **Anemometer**  
 Manufacturer : Holdpeak  
 Model : NA  
 Serial Number : **HP-866B-02**  
 Type : Digital  
 Range : 2 m/s to 30 m/s  
 Calibrated Range : 2 m/s to 30 m/s  
 Readability : 0.01 m/s  
 Accuracy : ±3%  
 As found : Good  
 Calibrated By : Hussein Alaa  
 Calibration Date : July 4, 2024  
**Calibration Due : 03 July 2025**



**ENVIRONMENTAL CONDITIONS DURING TEST**

Ambient Temperature : 26.0 ± 2°C Humidity : 40.0 ± 5 % RH Atm. Pressure 1013.0 ± 10 hPa

**CALIBRATION METHOD**

The above equipment has been calibrated in accordance with ASTM D5096-02

**REFERENCE EQUIPMENT USED :**

DESCRIPTION	SERIAL NO.	CERTIFICATE NO.	CALIBRATION DATE	DUE DATE
Benchtop Wind Tunnel	T225896 F	LT-22082811	25-Aug-23	24-Aug-24

**CALIBRATION TEST RESULTS**

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Deviation %
m/s	m/s	m/s	%
5.01	5.00	0.01	0.2
9.98	10.00	-0.02	-0.2
15.05	15.00	0.05	0.3
20.01	20.00	0.01	0.1
29.89	30.00	-0.11	-0.4

Acceptance Limit : ±3%

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
- The results are post adjustment.