

## Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 14, 2024

Page 1 of 1

WORK ORDER NUMBER	: By Contract	CUSTOMER DETAILS	
JOB NUMBER	: QC/JN/24/00278	Name	: Halliburton Worldwide IRAQ (CMT)
CERTIFICATE NUMBER	: QC240714-15	Address	: Western Burjesia, Oil Street, District Zubair-South Iraq

### EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Relief Valve  
Manufacturer : Kunkle Valves  
Model : NA  
Serial Number : HB-CMT-PRV-2084  
Set/Pressure : 36 psi  
Tolerance :  $\pm 2$  psi  
As Found : In Tolerance  
Calibration Date : July 14, 2024  
**Calibration Due** : 13 July 2026  
Last Calibration : NA



### ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature :  $35.0 \pm 2^{\circ}\text{C}$  Humidity :  $30.0 \pm 5\% \text{ RH}$  Atm. Pressure :  $1011.0 \pm 10 \text{ hPa}$

### CALIBRATION METHOD

Standard Method API 527 was used to test the unit under test. Test pressure was applied on inlet of unit under test and test was repeated 3 times. Identical repeatings were carried out as recommended in used standard.

### REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	1505119397	Sino Instruments / HS108	20 Nov, 2023	19 Nov, 2024

### CALIBRATION TEST RESULTS

Set Pressure on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
psi	psi	psi	$\pm$ psi
36	35.3	0.7	2.0

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- ☒ The results are as found (no adjustment done).  
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	