

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 18, 2024

Page 1 of 1

REQUEST NUMBER : By Contract	APPROVED BY: _____
JOB NUMBER : QC/JN/24/00279	_____
CERTIFICATE NUMBER : QC240718-02	_____



CUSTOMER DETAILS

Name : Hilong Oil Service & Engineering Company
 Address : Basra - Iraq
 Department : HL100

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
 Manufacturer : **Baoji Instruments**
 Model : 00000314
 Serial Number : **DL103**
 Location : MUD PUMP 2
 Type : Analog
 Range : 0 to 60 MPa
 Calibrated Range : 0 to 60 MPa
 Resolution : 1 MPa
 Accuracy : 2.5
 Tolerance : ±2.5 % of Full Scale
 As found : In Tolerance
 Calibrated By : Asjad Rafiq
 Calibration Date : July 18, 2024
 Calibration Due : 17 July 2027
 Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 25.0 ± 2°C Humidity : 44.0 ± 5 % RH Atm. Pressure 1011.0 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
MPa	MPa	MPa	%
0	0.00	0.00	0.0
15	15.08	-0.08	-0.1
30	30.11	-0.11	-0.2
45	45.13	-0.13	-0.2
60	60.18	-0.18	-0.3

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- The results are as found (no adjustment done).
 The results are post adjustment.

