

Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com



Page 1 of 3

Date of Issue: July 29, 2024

	:		APPROVED BY	LAB NORTH RUMAILAH
JOB NUMBER	:	QC/JN/24/00296		
CERTIFICATE NUMBER	:	QC240729-06		

CUSTOMER DETAILS

Name : Halliburton Worldwide IRAQ
Address : Western Burjesia, Oil Street, Zubair, South Iraq
Department : WPS

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Blaster's Multimeter
Type of Indication : Digital
Manufacturer : Thomas Instruments Inc.
Model : 109
Serial Number : FB17076
SAP Number : 300168451



Calibrated Range :

Voltage (DC)	0.1 mV	to	1500 V
Voltage (AC)	0.1 mV	to	1000 V
Current (DC)	0.1 μ A	to	2 A
Current (AC)	0.1 μ A	to	2 A
Resistance	0.1 Ω	to	20 M Ω

Resolution :

Voltage (DC)	0.1mV	0.0001V	0.001V	0.01V	0.1V	1V
Voltage (AC)	0.1mV	0.0001V	0.001V	0.01V	0.1V	1V
Current (DC)	0.1 μ A	0.001mA	0.01mA	0.1mA	1A	
Current (AC)	0.1 μ A	0.001mA	0.01mA	0.1mA	1A	
Resistance	0.1 ohm	0.001k ohm	0.01k ohm	0.1k ohm	0.001M ohm	0.01M Ohm

As Found Condition : Good
Calibrated By : Asjad Rafiq
Calibration Date : July 29, 2024
Calibration Due : July 28, 2026
Last Calibration Test : July 25, 2022

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22 $^{\circ}$ C $\pm$ 2 $^{\circ}$ C
Relative Humidity : 40 %RH $\pm$ 5 %RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with QC Calibration Procedure # QC/CP/E/01

The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

TRACEABILITY

The measurements made by Quality Control Labs, realize the physical units of measurements (SI), through its state of the art calibration standards that are controlled and maintained by QC.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Multifunction Calibrator	Fluke, USA	5522A	2806902	8/25/2023	8/24/2024
Ref Multimeter	Fluke, USA	8508A	276568089	8/25/2023	8/24/2024
Decade Resistance Box	Corpico	RBB6-B	18F-1093	8/25/2023	8/24/2024



Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138

CERTIFICATE OF CALIBRATION

REQUEST NUMBER:
JOB NUMBER:

By Mail
QC/JN/24/00296

CERTIFICATE NUMBER

QC240729-06
Page 2 of 3

CALIBRATION TEST RESULTS

Measurement Data for DC Voltage

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
mV	mV	mV	± (mV)
200.2	199.97	0.23	1
V	V	V	± (V)
2.008	2.000	0.0080	0.1
19.97	20.000	-0.0300	0.2
199.8	200.000	-0.2000	0.2
1000	1001.400	-1.4000	2.0
1500	1502.100	-2.1000	2.0

Measurement Data for AC Voltage @ 60 Hz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
m V	m V	m V	m V
199.9	199.95	-0.05	1
V	V	V	V
1.992	1.994	-0.0020	0.1
20.04	20.010	0.0300	0.1
200.2	199.95	0.2500	0.5
999	1000.02	-1.0230	4.0

Measurement Data for Resistance

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
Ω	Ω
0	0

Before Adjustment	After Adjustment
kΩ	kΩ
0	0

Before Adjustment	After Adjustment
MΩ	MΩ
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
Ω	Ω	Ω	± (Ω)
199.8	200.05830	-0.25830	0.10
kΩ	kΩ	kΩ	± (kΩ)
1.996	2.002100	-0.006100	0.10
19.99	20.00210	-0.012100	0.10
200.0	200.0878	-0.097800	1.00
MΩ	MΩ	MΩ	± (KΩ)
2.010	2.001000	0.009000	1.00
20.00	20.0142	-0.014200	1.00



Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138

CERTIFICATE OF CALIBRATION

REQUEST NUMBER: By Mail
JOB NUMBER: QC/JN/24/00296

CERTIFICATE NUMBER

QC240729-06

Page 3 of 3

Measurement Data for DC Current

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
mA	mA
0	0

Before Adjustment	After Adjustment
A	A
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μA	μA	μA	$\pm (\mu A)$
200.0	199.995	0.005	0.1
mA	mA	mA	$\pm (mA)$
200.020	200.035	-0.015	0.1
20.03	19.950	0.080	0.1
199.8	199.992	-0.192	1
A	A	A	$\pm (mA)$
2.0	1.990000	0.010000	1

Measurement Data for AC Current @ 60 Hz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
mA	mA
0	0

Before Adjustment	After Adjustment
A	A
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μA	μA	μA	$\pm (\mu A)$
200.0	199.996	0.004	0
mA	mA	mA	$\pm (mA)$
1.997	1.993	0.004	0.1
20.00	19.993	0.007	0.1
200.0	199.982	-0.004	1
A	A	A	$\pm (A)$
2.0	1.997200	0.002800	1

Results :


Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	ALI Talib HB48903 Date: 30/07/2024 Signature  Haliburton



Address : North Rumaila, Al Takamul Yard
Contact : +9647810009138