

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 29, 2024

Page 1 of 1

REQUEST NUMBER :	Email	APPROVED BY LAB INCHARGE QC
JOB NUMBER :	QC/JN/24/00296	
CERTIFICATE NUMBER :	QC240729-21	Asjad Rafiq

CUSTOMER DETAILS

Name : **Halliburton Worldwide IRAQ**
 Address : Western Burjesia, Oil Street, District Zubair-South Iraq
 Department : WPS

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **3/4" SUCKER ROD NOT GO PIN THREAD RING**
 Manufacturer : GAGE ASSEMBLY
 Model : UNS-2A-P6
 SAP. Number : NA
 Serial Number : 4
 Range : 0.9683 Inches
 Tolerance : +0.0006 Inches
 As Found : In Tolerance
 Calibration Date : July 29, 2024
Calibration Due : July 28, 2025
 Last Calibration : June 19, 2023



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : 22.0 ± 2°C Humidity : 15.0 ± 5 % RH Atm. Pressure 1013.6 ± 10 hPa

CALIBRATION METHOD

The above equipment has been calibrated in accordance with API-SPEC-11B

1. Check That Ring Gauge Is Clearly Identified With Serial Numbers
2. Clean Ring Gauge Until It's Free Of All Visible Contamination.
5. Verify The Ring Gauge inside Diameter In 6 Places By Internal Micrometer
6. Record Measurement Taken On Calibration Report.

REFERENCE EQUIPMENT USED :

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Vernier Caliper	07414025	Mitutoyo / CD-12CX	15-Mar-24	14-Sep-24
Internal Micrometer	58622147	Mitutoyo / 345-350-30	15-Mar-24	14-Sep-24

CALIBRATION TEST RESULTS

Actual Diameter Readings For Ring Gauge diameter (Inches)						Average
R1	R2	R3	R4	R5	R6	Inches
0.9684	0.9684	0.9685	0.9685	0.9684	0.9685	0.9684

Results :

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	ALI Talib HB48903 Date: 29-07-2024 Signature  Haliburton