

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 30, 2024

REQUEST NUMBER : By Contract
JOB NUMBER : QC/JN/24/00302
CERTIFICATE NUMBER : QC240730-24

APPROVED BY: ALI CHANGI QC



CUSTOMER DETAILS

Name : slb Basra
Address : North Rumaila Base south Iraq
Department : Slickline

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : Pressure Relief Valve
Manufacturer : NA
Model : NA
Serial Number : F03
Location : Compressor
Set/Pressure : 115 psi
Tolerance : $\pm 3\%$
Calibrated By : Hussein Alaa
Calibration Date : July 30, 2024
Calibration Due : July 29, 2025



ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : $25.0 \pm 1^\circ\text{C}$ Humidity : $44.0 \pm 5\% \text{ RH}$ Atm. Pressure : $1011.0 \pm 10 \text{ hPa}$

CALIBRATION METHOD

Standard Method API 527 was used to test the unit under test. Test pressure was applied on inlet of unit under test and test was repeated 3 times. Identical repeatings were carried out as recommended in used standard.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	1505119397	Sino Instruments / HS108	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
psi	psi	psi	$\pm$ psi
115	119	-4	3.45

Results :

The above Value Indicates That the Instrument complies with the Specified Maximum Permissible error tolerance.

At The Time of Calibration. and ready for use.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- ☒ The results relate to measurements taken at nominal value only.
- ☐ Adjustments were made prior to calibration.
- ☐ Accessories were fitted.

~ END OF DOCUMENT ~