

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq


• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: July 30, 2024

Page 1 of 1

REQUEST NUMBER : By Contract
JOB NUMBER : QC/JN/24/00302
CERTIFICATE NUMBER : QC240730-32

APPROVED BY: [Signature]



CUSTOMER DETAILS

Name : slb Basra
Address : North Rumaila Base south Iraq
Department : Slickline

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **Pressure Gauge**
Manufacturer : McDanile
Model : 316L
Serial Number : 153551127
Location : NA
Type : Analog
Range : 0 to 15000 psi
Calibrated Range : 0 to 15000 psi
Resolution : 200 psi
Tolerance : ± 1 % of Full Scale
As found : In Tolerance
Calibrated By : Hussein Alaa
Calibration Date : July 30, 2024
Calibration Due : 29 July 2025
Last Calibration Date : NA



DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

1. No adjustments were made prior to calibration.
2. Overload test was not performed.
3. No accessories were fitted.
4. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature : $25.0 \pm 2^{\circ}\text{C}$ Humidity : $44.0 \pm 5\% \text{ RH}$ Atm. Pressure : $1011.0 \pm 10 \text{ hPa}$

CALIBRATION METHOD

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Error Percentage
psi	psi	psi	$\pm$
0	0	0	0.0
5000	4975	25.00	0.2
7500	7483	17.00	0.1
10000	9991	9.00	0.1
15000	14993	7.00	0.0

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.