


Calibration Certificate

Al Takamul Yard North Rumailah, Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq

Date of Issue: September 2, 2024

Page 1 of 5

REQUEST NUMBER : By Mail	APPROVED BY:  CHARGE QC
JOB NUMBER : QC/JN/24/00327	
CERTIFICATE NUMBER : QC240902-01	

CUSTOMER DETAILS

Name : **Egyptian Maintenance Company(EMC) IRAQ**
 Address : Iraq, Basra, North Rumaila

EQUIPMENT IDENTIFICATION AND SPECIFICATIONS

Description : **True RMS Multimeter**
 Type of Indication : Digital
 Manufacturer : FLUKE
 Model : 87V
 Asset No. : N/A
 Serial Number : **23610226**



Calibrated Range:

Voltage (DC)	0.01 mV	to	1000 V
Voltage (AC) @ 50 Hz	0.1 mV	to	1000 V
Voltage (AC) @ 1 kHz	0.1 mV	to	1000 V
Current (DC)	0.01 μA	to	10 A
Current (AC) @ 50 Hz	0.1 μA	to	10 A
Current (AC) @ 1 kHz	0.1 μA	to	10 A
Resistance	1 Ω	to	50 M Ω

Resolution:

Voltage (DC)	0.01 mV	0.001 mV	0.001 V
Voltage (AC) @ 50 Hz	0.1 mV	0.001 mV	0.001 V
Voltage (AC) @ 1 kHz	0.1 mV	0.001 mV	0.001 V
Current (DC)	0.01 μA	0.001 mA	0.001 A
Current (AC) @ 50 Hz	0.1 μA	0.001 mA	0.001 A
Current (AC) @ 1 kHz	0.1 μA	0.001 mA	0.001 A
Resistance	0.1 Ω	0.001 K Ω	0.001 M Ω

Location : QC Labs North Rumailah, Iraq
 Calibrated By : Asjad Rafiq
Calibration Date : September 2, 2024
 Calibration Due : **Recommended Validity : one (1) year from the date of calibration, (Where Required) September 1, 2025**

ENVIRONMENTAL CONDITIONS DURING TEST

Ambient Temperature	:	22 °C	±	2 °C
Relative Humidity	:	45 %RH	±	5 %RH

CALIBRATION METHOD

The above equipment has been calibrated in accordance with QC Calibration Procedure # QC/CP/E/01
 The deviations of the measurements obtained from UUC with respect to reference standards are determined to obtain the error.

REFERENCE EQUIPMENT USED :

DESCRIPTION	MAKE	MODEL #	SERIAL #	CALIBRATION DATE	CALIBRATION DUE DATE
Multifunction Calibrator	Fluke, USA	5522A	2806902	8/22/2024	8/21/2025
Ref Multimeter	Fluke, USA	8508A	276568089	8/22/2024	8/21/2025
Decade Resistance Box	Corpico	RBB6-B	18F-1093	8/22/2024	8/21/2025

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

REQUEST NUMBER: By Email
JOB NUMBER: QC/JN/24/00327

QC240902-01

Page 2 of 5

CALIBRATION TEST RESULTS

Measurement Data for DC Voltage

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μV	μV
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μV	μV	μV	$\pm (\mu\text{V})$
99.9	100.00	-0.10	0.1
mV	mV	mV	$\pm (\text{mV})$
0.9	1.00	-0.10	0.1
9.8	10.00	-0.20	0.1
100	100.00	0.00	0.1
500	500.00	0.00	0.1
600	601.23	-1.23	0.1
V	V	V	$\pm (\text{V})$
1.002	1.00	0.00	0.1
-1.002	-1.00	0.00	0.1
10.01	10.00	0.01	0.1
100	100.00	0.00	0.1
300	300.00	0.00	0.1
500	500.00	0.00	1
901	900.00	1.00	1
1000	1000.00	0.00	1

Measurement Data for AC Voltage @ 50 Hz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μV	μV
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μV	μV	μV	$\pm (\mu\text{V})$
99.9	100.000	-0.100	0.1
mV	mV	mV	$\pm (\text{mV})$
0.9	1.000	-0.100	0.1
9.8	10.000	-0.200	0.1
100	100.000	0.000	0.1
500	499.998	0.002	0.1
600	598.650	1.350	0.1
V	V	V	$\pm (\text{V})$
1.002	0.999980	0.002020	0.1
10.01	9.99998	0.01002	0.1
100.0	100.0130	-0.0130	0.1
300.0	300.013	-0.013	0.1
500	500.013	-0.013	1
901	900.080	0.920	1
1000	1000.000	0.000	1

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC240902-01

Page 3 of 5

REQUEST NUMBER: By Email
JOB NUMBER: QC/JN/24/00327

Measurement Data for AC Voltage @ 1 kHz
Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μV	μV
0	0

Before Adjustment	After Adjustment
mV	mV
0	0

Before Adjustment	After Adjustment
V	V
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μV	μV	μV	$\pm (\mu\text{V})$
99.9	100.00	-0.10	0.1
mV	mV	mV	$\pm (\text{mV})$
0.9	1.001	-0.101	0.1
9.8	10.001	-0.201	0.1
100	100.001	-0.001	0.1
500	500.000	0.000	0.1
600	599.950	0.050	0.1
V	V	V	$\pm (\text{V})$
1.002	1.000010	0.001990	0.1
10.01	10.00001	0.00999	0.1
100.0	99.9870	0.0130	0.1
300.0	299.987	0.013	0.1
500	499.987	0.013	1
901	899.930	1.070	1
1000	999.989	0.011	1

Measurement Data for DC Current
Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μA	μA	μA	$\pm (\mu\text{A})$
99.9	99.999	-0.099	0.1
mA	mA	mA	$\pm (\text{mA})$
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	99.999	0.001	0.1
500	499.994	0.006	0.1
600	600.125	-0.125	0.1
A	A	A	$\pm (\text{A})$
1.002	1.000010	0.001990	0.01
-1.002	-1.000030	-0.001970	0.01
3.002	2.999950	0.002050	0.01
5.003	4.999950	0.003050	0.06
11.003	10.999900	0.003100	0.06

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC240902-01

Page 4 of 5

REQUEST NUMBER: By Email
JOB NUMBER: QC/JN/24/00327

Measurement Data for AC Current @ 50 Hz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μA	μA	μA	$\pm (\mu\text{A})$
99.9	99.999	-0.099	0.1
mA	mA	mA	$\pm (\text{mA})$
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	100.000	0.000	0.1
500	500.030	-0.030	0.1
900	899.900	0.100	0.2
A	A	A	$\pm (\text{A})$
1.002	1.000100	0.001900	0.01
3.002	3.000040	0.001960	0.01
5.003	5.000040	0.002960	0.06
9.9	10.235000	-0.335000	0.06

Measurement Data for AC Current @ 1 kHz

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Before Adjustment	After Adjustment
μA	μA
0	0

Measurement Data for AC Current @ 1 kHz

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
μA	μA	μA	$\pm (\mu\text{A})$
99.9	99.999	-0.099	0.1
mA	mA	mA	$\pm (\text{mA})$
0.9	1.000	-0.100	0
9.8	10.000	-0.200	0.1
100	100.020	-0.020	0.1
500	500.040	-0.040	0.1
600	601.230	-1.230	0.2
A	A	A	$\pm (\text{A})$
1.002	0.999950	0.002050	0.01
3.002	2.999780	0.002220	0.01
5.003	4.999780	0.003220	0.06
10.002	10.002568	-0.000568	0.06

CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER

QC240902-01

Page 5 of 5

REQUEST NUMBER: By Email
JOB NUMBER: QC/JN/24/00327

Measurement Data for Resistance

Zero or Offset Readings of UUC

Before Adjustment	After Adjustment
Ω	Ω
0	0

Before Adjustment	After Adjustment
$k\Omega$	$k\Omega$
0	0

Before Adjustment	After Adjustment
$M\Omega$	$M\Omega$
0	0

Readings on UUC	Readings on Ref. Standard	Error	Uncertainty (95 % C.L.)
Ω	Ω	Ω	$\pm (\Omega)$
1.1	1.00010	0.09990	0.08
10.1	9.99990	0.10010	0.10
100.1	99.9994	0.1006	0.08
900.2	900.0000	0.2000	0.08
$k\Omega$	$k\Omega$	$k\Omega$	$\pm (k\Omega)$
1.002	1.000000	0.002000	0.01
10.002	9.99996	0.00204	0.01
100.1	100.0006	0.0994	0.06
500.1	499.9993	0.1007	0.05
900.2	899.999	0.2010	0.08
$M\Omega$	$M\Omega$	$M\Omega$	$\pm (M\Omega)$
1.001	1.000001	0.000999	0.12
10.000	10.00003	-0.00003	0.01
30.002	29.9990	0.0030	0.02
50.002	49.9990	0.0030	0.02

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. The uncertainty evaluation has been carried out in accordance with international practice.

DEVIATION FROM STANDARD METHOD : None

REMARK (S) :

- The results are as found (no adjustment done).
- The results are post adjustment.

~ END OF DOCUMENT ~

This certificate is issued in accordance with the laboratory accreditation requirements as per ISO/IEC17025:2017.

This certificate may not be reproduced other than in full, unless permission for the publication of an approved extract has been obtained in writing from the director of QC. It does not of itself impute to the subject of the calibration any attributes beyond those shown by the data contained herein.