

Certificate of Calibration

Al Takamul Yard, North Rumailah South Iraq

• Phone : +964 7810009138 • www.qualitycontrol-iraq.com • E-mail: op@qualitycontrol-iraq.com

Date of Issue: September 9, 2024**Page 1 of 1**

REQUEST NUMBER	: By Contract	CUSTOMER DETAILS	
JOB NUMBER	: QC/JN/24/00346	Name	Halliburton Worldwide IRAQ (WPS)
CERTIFICATE NUMBER	: QC240909-01	Address	Western Burjesia, Oil Street, District Zubair-South Iraq

Description : **Pressure Gauge**
Manufacturer : NA
Model : NA
Serial Number : **PG24910**
Type : Analog
Range : 0 to 350 psi
Calibrated Range : 0 to 350 psi
Resolution : 10 psi
Accuracy : NA
Tolerance : ± 1.0 % of Full Scale
As found : In Tolerance
Calibration Date : September 9, 2024
Calibration Due : **08 September 2025**
Last Calibration : NA

**DEVIATION FROM STANDARD METHOD** : None**REMARK (S) :**

1. No adjustments were made prior to calibration.
2. No accessories were fitted.
3. Values were rounded in computation.

ENVIRONMENTAL CONDITIONS DURING TESTAmbient Temperature : 22.0 $\pm$ 2°C Humidity : 15.0 $\pm$ 5 % RH Atm. Pressure 1013.0 $\pm$ 10 hPa**CALIBRATION METHOD**

The above equipment has been calibrated in accordance with Calibration Procedure QC/CP/R/03 Rev. 01

The above instrument has been calibrated in increasing and decreasing pressure mode using a Digital Pressure Calibrator as the reference standard having traceability to Internationally recognised standards.

REFERENCE EQUIPMENT USED

DESCRIPTION	SERIAL NO.	MAKE / MODEL	CALIBRATION DATE	CALIBRATION DUE DATE
Digital Pressure Gauge	468802	AMETEK / XP2i	20 Nov, 2023	19 Nov, 2024

CALIBRATION TEST RESULTS

Measured Values on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Full Scale Error %
psi	psi	psi	%
0	0.00	0.00	0.0
90	91.2	-1.20	-0.3
180	181.6	-1.60	-0.5
270	271.5	-1.50	-0.4
350	351.8	-1.80	-0.5

Calibration results were found to conform as per specified accuracy requirements. Above Instrument has **PASSED** its Calibration.

- ☒ The results are as found (no adjustment done).
☐ The results are post adjustment.

CALIBRATED BY	REVIEWED & APPROVED BY LAB INCHARGE	CLIENT
 Hussein Alaa	 LAB INCHARGE Asjad Rafiq	